

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134418

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3158

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	01/27/2025	09:15
2	Q1165-01		1	13	100.00	>212.0	>212.0	01/27/2025	09:45
3	Q1167-02		1	11	100.00	>212.0	>212.0	01/27/2025	10:15
4	Q1167-02DUP		1	11	100.00	>212.0	>212.0	01/27/2025	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66134418

WorkList Name : fp-1-27

WorkList ID : 187162

Department : Wet-Chemistry

Date : 01-27-2025 08:21:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1165-01	91624-F	Water	Flash Point	Cool 4 deg C	PSEG03	N41	01/23/2025	1010B
Q1167-02	34784	Water	Flash Point	Cool 4 deg C	PSEG03	N41	01/23/2025	1010B

Date/Time 01/27/2025 08:30
 Raw Sample Received by: RM
 Raw Sample Relinquished by: [Signature]

Date/Time 01/27/2025 10:50
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]