



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 1/28/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 14:20
In Date: 01/27/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 01/28/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB134420

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q1190-01	A44S2	1	1.15	8.81	9.96	6.65	62.4	
Q1190-02	A44S3	2	1.16	8.54	9.7	8.18	82.2	
Q1190-03	A44S4	3	1.16	8.40	9.56	8.13	83.0	
Q1190-04	A44S5	4	1.15	8.67	9.82	7.53	73.6	
Q1190-05	A44S6	5	1.15	8.81	9.96	6.61	62.0	
Q1190-06	A44S6MS	6	1.15	8.81	9.96	6.61	62.0	
Q1190-07	A44S6MSD	7	1.15	8.81	9.96	6.61	62.0	
Q1190-08	A44S7	8	1.12	8.72	9.84	7.92	78.0	
Q1190-09	A44S8	9	1.17	8.82	9.99	7.6	72.9	
Q1190-10	A44S9	10	1.17	8.60	9.77	7.71	76.0	
Q1190-11	A44T0	11	1.15	8.44	9.59	7.15	71.1	
Q1190-12	A44T1	12	1.18	8.69	9.87	7.13	68.5	
Q1190-13	A44T2	13	1.16	8.48	9.64	8.91	91.4	
Q1190-14	A44T3	14	1.15	8.82	9.97	8.56	84.0	
Q1190-15	A44T4	15	1.16	8.82	9.98	8.98	88.7	
Q1190-16	A44T6	16	1.17	8.65	9.82	8.88	89.1	
Q1190-17	A44T7	17	1.16	8.65	9.81	7.95	78.5	
Q1190-18	A44X1	18	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
Q1190-19	A44X3	19	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
Q1190-20	A44Y2	20	1.00	1.00	2.00	2.00	100.0	FB solids
Q1190-21	A44Y3	21	1.00	1.00	2.00	2.00	100.0	FB solids
Q1190-22	VHBLK001	22	1.00	1.00	2.00	2.00	100.0	vhblk

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$