

LB134479


7470A

INSTRUMENT ID: CV3

Linear

 μ Abs.:

40814



A= 0.0000e+000

B= 2.4123e-004

C= 4.5607e-002

Rho= 0.9994939

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.057	0.057	47	0.000	47	0				
0.2	0.200	0.207	0.007	667	0.0 %	667					
2.5	2.500	2.304	-0.196	9362	0.0 %	9362					
5.0	5.000	5.126	0.126	21062	0.0 %	21062					
7.5	7.500	7.615	0.115	31377	0.0 %	31377					
10.0	10.000	9.891	-0.109	40814	0.0 %	40814					

94.0
1230031

LB134479 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Jan 2025 09:25:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.0000	7470A	PPB	30 Jan 2025 09:28:32	S	Std
0.2 - 1	50.2	667	-	0.2000	7470A	PPB	30 Jan 2025 09:30:49	S	Std
2.5 - 1	52.5	9362	-	2.5000	7470A	PPB	30 Jan 2025 09:33:07	S	Std
5.0 - 1	55	21062	-	5.0000	7470A	PPB	30 Jan 2025 09:38:00	S	Std
7.5 - 1	57.5	31377	-	7.5000	7470A	PPB	30 Jan 2025 09:42:57	S	Std
10.0 - 1	59.6	40814	-	10.0000	7470A	PPB	30 Jan 2025 09:45:14	S	Std
ICV75 - 1	ICV75	16880	4.1176	-7470A	PPB	30 Jan 2025 09:48:06	U	SMPL	
ICB75 - 1	ICB75	-195	-0.0014	-7470A	PPB	30 Jan 2025 09:50:21	U	SMPL	
CCV87 - 1	CCV87	21655	5.2695	-7470A	PPB	30 Jan 2025 09:52:37	U	SMPL	
CCB87 - 1	CCB87	-212	-0.0055	-7470A	PPB	30 Jan 2025 09:54:53	U	SMPL	
CRA - 1	CRA	614	0.1937	-7470A	PPB	30 Jan 2025 09:59:42	U	SMPL	
HighStd - 1	HighStd	41852	10.1417	-7470A	PPB	30 Jan 2025 10:01:56	U	SMPL	
ChkStd - 1	ChkStd	29903	7.2592	-7470A	PPB	30 Jan 2025 10:04:11	U	SMPL	
PB166355BL - 1	PBW	111	0.0724	-7470A	PPB	30 Jan 2025 10:11:54	U	SMPL	
PB166355BS - 1	LCSW	18526	4.5147	-7470A	PPB	30 Jan 2025 10:14:09	U	SMPL	
Q1205-02 - 1	VNJ-236	45	0.0565	-7470A	PPB	30 Jan 2025 10:16:25	U	SMPL	
Q1206-04 - 1	JPP-20.1-012725	66	0.0615	-7470A	PPB	30 Jan 2025 10:18:43	U	SMPL	
Q1206-08 - 1	JPP-16.3-012725	47	0.0569	-7470A	PPB	30 Jan 2025 10:21:00	U	SMPL	
Q1207-04 - 1	JPP-2.1-012725	76	0.0639	-7470A	PPB	30 Jan 2025 10:23:17	U	SMPL	
Q1207-08 - 1	JPP-5.1-012725	132	0.0774	-7470A	PPB	30 Jan 2025 10:25:33	U	SMPL	
CCV88 - 1	CCV88	21976	5.3469	-7470A	PPB	30 Jan 2025 10:27:48	U	SMPL	
CCB88 - 1	CCB88	-207	-0.0043	-7470A	PPB	30 Jan 2025 10:30:04	U	SMPL	
Q1207-12 - 1	JPP-4.5-012725	79	0.0647	-7470A	PPB	30 Jan 2025 10:32:22	U	SMPL	
Q1207-16 - 1	JPP-16.2-012725	62	0.0606	-7470A	PPB	30 Jan 2025 10:34:37	U	SMPL	
Q1207-20 - 1	JPP-20.2-012725	20	0.0504	-7470A	PPB	30 Jan 2025 10:36:52	U	SMPL	
Q1209-04 - 1	WC-4	37	0.0545	-7470A	PPB	30 Jan 2025 10:39:08	U	SMPL	
Q1209-08 - 1	WC-5	16	0.0495	-7470A	PPB	30 Jan 2025 10:41:25	U	SMPL	
Q1209-08DUP - 1	WC-5DUP	-30	0.0384	-7470A	PPB	30 Jan 2025 10:43:41	U	SMPL	
Q1209-08MS - 1	WC-5MS	17702	4.3159	-7470A	PPB	30 Jan 2025 10:45:58	U	SMPL	
Q1209-08MSD - 1	WC-5MSD	17040	4.1562	-7470A	PPB	30 Jan 2025 10:48:15	U	SMPL	
CCV89 - 1	CCV89	20705	5.0403	-7470A	PPB	30 Jan 2025 10:53:13	U	SMPL	
CCB89 - 1	CCB89	94	0.0683	-7470A	PPB	30 Jan 2025 10:58:06	U	SMPL	
PB166318TB - 1	PB166318TB	89	0.0671	-7470A	PPB	30 Jan 2025 11:03:37	U	SMPL	
Q1209-08LX5 - 1		54	0.0586	-7470A	PPB	30 Jan 2025 11:05:52	U	SMPL	
Q1209-08A - 1		17633	4.2993	-7470A	PPB	30 Jan 2025 11:08:07	U	SMPL	
CCV90 - 1	CCV90	20812	5.0661	-7470A	PPB	30 Jan 2025 11:15:11	U	SMPL	
CCB90 - 1	CCB90	-44	0.0350	-7470A	PPB	30 Jan 2025 11:17:27	U	SMPL	