

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Niha

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB134486

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP111700
Residual chlorine ICV-LCS, 0.4PPM	WP111701
Chlorine Calibration std, 0.1ppm	WP111705
Chlorine Calibration std, 0.2ppm	WP111704
Chlorine Calibration std, 0.8ppm	WP111703
Chlorine Calibration std, 0.0ppm	WP111696
Chlorine Calibration std, 1.6ppm	WP111706
Residual Chlorine LOQ 0.1ppm	WP111699
Residual Chlorine Calibration and CCV std, 0	WP111702
Total Chlorine Powder Pillows	W3147

Intercept: 0.0186

Slope: 0.9286

Regression: 0.999725

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.02		01/30/2025	13:00
2	CAL2	0.1	1	0.000	0.120	0.120	0.11	9	01/30/2025	13:03
3	CAL3	0.2	1	0.000	0.200	0.200	0.20	-2.5	01/30/2025	13:06
4	CAL4	0.4	1	0.000	0.410	0.410	0.42	5.2	01/30/2025	13:09
5	CAL5	0.8	1	0.000	0.760	0.760	0.80	-0.3	01/30/2025	13:12
6	CAL6	1.6	1	0.000	1.500	1.500	1.60	-0.3	01/30/2025	13:15



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Niha

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB134486

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4000	0.4000	0.4110	01/30/2025	13:18
2	ICB				1	0.0000	0.0000	0.0000	-0.0200	01/30/2025	13:21
3	CCV1			0.4	1	0.0000	0.3900	0.3900	0.4000	01/30/2025	13:24
4	CCB1				1	0.0000	0.0000	0.0000	-0.0200	01/30/2025	13:27
5	LB134486BL	50	50		1	0.0000	0.0000	0.0000	-0.0200	01/30/2025	13:30
6	LB134486BS	50	50	0.4	1	0.0000	0.4100	0.4100	0.4210	01/30/2025	13:33
7	Q1168-07	50	50		1	0.0000	0.0400	0.0400	0.0230	01/30/2025	13:36
8	Q1168-08	50	50		1	0.0000	0.1100	0.1100	0.0980	01/30/2025	13:39
9	CCV2			0.4	1	0.0000	0.3900	0.3900	0.4000	01/30/2025	13:42
10	CCB2				1	0.0000	0.0000	0.0000	-0.0200	01/30/2025	13:45