

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134492

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/30/2025	16:50
2	CAL2	1	Water	NA	NA	20.1	6.99	01/30/2025	16:51
3	CAL3	1	Water	NA	NA	20.1	10.02	01/30/2025	16:53
4	ICV	1	Water	NA	NA	20.3	7.01	01/30/2025	16:55
5	CCV1	1	Water	NA	NA	20.2	2.01	01/30/2025	17:00
6	Q1234-02	1	Water	NA	NA	20.8	6.85	01/30/2025	17:10
7	Q1234-02DUP	1	Water	NA	NA	20.9	6.86	01/30/2025	17:11
8	Q1234-03	1	Water	NA	NA	20.8	7.01	01/30/2025	17:15
9	Q1234-04	1	Water	NA	NA	20.9	7.02	01/30/2025	17:20
10	Q1234-05	1	Water	NA	NA	20.2	6.02	01/30/2025	17:22
11	Q1234-06	1	Water	NA	NA	20.7	6.85	01/30/2025	17:25
12	Q1234-07	1	Water	NA	NA	20.1	6.83	01/30/2025	17:30
13	CCV2	1	Water	NA	NA	20.3	12.02	01/30/2025	17:35