

LB134514


SFAM01.1

INSTRUMENT ID: (V)

Linear

 μ Abs.:

52137



A= 0.0000e+000

B= 1.9235e-004 slope

C= 1.6048e-002 y-intercept

Rho= 0.9997877

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/cd
0.00	0.000	0.050	0.050	174	0.000	174					1
0.05	0.050										3
0.20	0.200	0.205	0.005	982	0.0 %	982					5
2.50	2.500	2.397	-0.103	12380	0.0 %	12380					-4
5.00	5.000	5.101	0.101	26435	0.0 %	26435					2
7.50	7.500	7.403	-0.097	38401	0.0 %	38401					1
10.0	10.000	10.045	0.045	52137	0.0 %	52137					0

LB134514

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	174	-		0 SFAM01.1	PPB	1/31/2025 16:43	Std
	0.2 S01	982	-		0.2 SFAM01.1	PPB	1/31/2025 16:46	Std
	2.5 S02	12380	-		2.5 SFAM01.1	PPB	1/31/2025 16:48	Std
	5 S03	26435	-		5 SFAM01.1	PPB	1/31/2025 16:50	Std
	7.5 S04	38401	-		7.5 SFAM01.1	PPB	1/31/2025 16:52	Std
	10 S05	52137	-		10 SFAM01.1	PPB	1/31/2025 16:55	Std
ICV078	ICV078	23075	4.4546 -		SFAM01.1	PPB	1/31/2025 17:00	SMPL
ICB078	ICB078	41	0.0239 -		SFAM01.1	PPB	1/31/2025 17:02	SMPL
CCV001	CCV001	25946	5.0068 -		SFAM01.1	PPB	1/31/2025 17:05	SMPL
CCB001	CCB001	1	0.0162 -		SFAM01.1	PPB	1/31/2025 17:07	SMPL
Q1223-01	A6301	5958	1.1621 -		SFAM01.1	PPB	1/31/2025 17:11	SMPL
Q1223-02	A6302	3948	0.7755 -		SFAM01.1	PPB	1/31/2025 17:14	SMPL
Q1223-03	A6303	955	0.1997 -		SFAM01.1	PPB	1/31/2025 17:16	SMPL
Q1223-04	A6304	1164	0.2399 -		SFAM01.1	PPB	1/31/2025 17:18	SMPL
Q1223-05	A6305	4331	0.8491 -		SFAM01.1	PPB	1/31/2025 17:21	SMPL
Q1223-06	A6305D	3979	0.7814 -		SFAM01.1	PPB	1/31/2025 17:23	SMPL
Q1223-07	A6305S	15712	3.0383 -		SFAM01.1	PPB	1/31/2025 17:25	SMPL
Q1223-08	A6306	1883	0.3782 -		SFAM01.1	PPB	1/31/2025 17:27	SMPL
Q1223-09	A6307	6207	1.21 -		SFAM01.1	PPB	1/31/2025 17:30	SMPL
Q1223-10	A6308	8899	1.7278 -		SFAM01.1	PPB	1/31/2025 17:32	SMPL
Q1223-11	A6309	353	0.0839 -		SFAM01.1	PPB	1/31/2025 17:34	SMPL
Q1223-12	A6320	609	0.1332 -		SFAM01.1	PPB	1/31/2025 17:36	SMPL
Q1223-13	A6321	1511	0.3067 -		SFAM01.1	PPB	1/31/2025 17:39	SMPL
Q1223-14	A6322	4503	0.8822 -		SFAM01.1	PPB	1/31/2025 17:41	SMPL
Q1223-15	A6323	5339	1.043 -		SFAM01.1	PPB	1/31/2025 17:43	SMPL
Q1223-16	A6324	4214	0.8266 -		SFAM01.1	PPB	1/31/2025 17:45	SMPL
PB166456BL	PBW456	12	0.0184 -		SFAM01.1	PPB	1/31/2025 17:48	SMPL
CCV002	CCV002	25435	4.9085 -		SFAM01.1	PPB	1/31/2025 17:50	SMPL
CCB002	CCB002	-4	0.0153 -		SFAM01.1	PPB	1/31/2025 17:52	SMPL