

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB134515

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/31/2025	16:45
2	CAL2	1	Water	NA	NA	20.1	7.00	01/31/2025	16:46
3	CAL3	1	Water	NA	NA	20.1	10.02	01/31/2025	16:47
4	ICV	1	Water	NA	NA	20.2	7.02	01/31/2025	16:50
5	CCV1	1	Water	NA	NA	20.2	2.01	01/31/2025	16:55
6	Q1241-04	1	Solid	20.02	20	20.5	8.62	01/31/2025	17:00
7	Q1241-04DUP	1	Solid	20.03	20	20.6	8.63	01/31/2025	17:01
8	Q1241-08	1	Solid	20.04	20	20.5	10.86	01/31/2025	17:05
9	Q1241-12	1	Solid	20.02	20	20.6	11.32	01/31/2025	17:15
10	Q1241-16	1	Solid	20.03	20	21.1	10.02	01/31/2025	17:25
11	Q1241-20	1	Solid	20.03	20	21.2	9.02	01/31/2025	17:33
12	Q1242-04	1	Solid	20.02	20	21.0	11.57	01/31/2025	18:00
13	Q1262-05	1	Solid	20.01	20	22.3	5.94	01/31/2025	18:05
14	CCV2	1	Water	NA	NA	20.3	12.02	01/31/2025	18:15

WORKLIST(Hardcopy Internal Chain)

134515

WorkList Name : corrosivity Q1242

WorkList ID : 187351

Department : Wet-Chemistry

Date : 01-31-2025 10:27:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1241-04	JPP-3.5-013025	Solid	Corrosivity	Cool 4 deg C	RUTW01	E11	01/30/2025	9045D
Q1241-08	JPP-5.3-013025	Solid	Corrosivity	Cool 4 deg C	RUTW01	E11	01/30/2025	9045D
Q1241-12	JPP-5.2-013025	Solid	Corrosivity	Cool 4 deg C	RUTW01	E11	01/30/2025	9045D
Q1241-16	JPP-5.4-013025	Solid	Corrosivity	Cool 4 deg C	RUTW01	E11	01/30/2025	9045D
Q1241-20	JPP-51.4-013025	Solid	Corrosivity	Cool 4 deg C	RUTW01	E11	01/30/2025	9045D
Q1242-04	JPP-6.2-013025	Solid	Corrosivity	Cool 4 deg C	RUTW01	E11	01/30/2025	9045D
Q1262-05	3762	Solid	Corrosivity	Cool 4 deg C	PSEG03	N41	01/31/2025	9045D

Date/Time 01/31/25 16:40
 Raw Sample Received by: Sh Cobb
 Raw Sample Relinquished by: RS (EST-156)

Date/Time 01/31/25 18:30
 Raw Sample Received by: RS (EST-156)
 Raw Sample Relinquished by: Sh Cobb