

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134519

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/01/2025	10:00
2	CAL2	1	Water	NA	NA	20.2	7.00	02/01/2025	10:01
3	CAL3	1	Water	NA	NA	20.2	10.03	02/01/2025	10:05
4	ICV	1	Water	NA	NA	20.2	6.99	02/01/2025	10:10
5	CCV1	1	Water	NA	NA	20.2	2.01	02/01/2025	10:14
6	Q1266-01	1	Water	NA	NA	20.2	7.73	02/01/2025	10:20
7	Q1266-01DUP	1	Water	NA	NA	20.3	7.74	02/01/2025	10:21
8	Q1266-02	1	Water	NA	NA	20.2	8.31	02/01/2025	10:27
9	Q1266-03	1	Water	NA	NA	20.4	8.50	02/01/2025	10:33
10	CCV2	1	Water	NA	NA	20.3	12.02	02/01/2025	10:37

WORKLIST(Hardcopy Internal Chain)

134519

WorkList Name : PH W Q1266 WorkList ID : 187389 Department : Wet-Chemistry Date : 02-01-2025 09:45:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1266-01	TRE-25-0004	Water	pH	Cool 4 deg C	PSEG03	N41	01/31/2025	9040C
Q1266-02	TRE-25-0010	Water	pH	Cool 4 deg C	PSEG03	N41	01/31/2025	9040C
Q1266-03	TRE-25-0012	Water	pH	Cool 4 deg C	PSEG03	N41	01/31/2025	9040C

Date/Time 02/01/25 09:50
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: RJ CEXT-146

Date/Time 02/01/25 13:00
 Raw Sample Received by: RJ CEXT-146
 Raw Sample Relinquished by: [Signature]