

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134525

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/03/2025	10:40
2	CAL2	1	Water	NA	NA	20.2	7.01	02/03/2025	10:41
3	CAL3	1	Water	NA	NA	20.3	10.02	02/03/2025	10:45
4	ICV	1	Water	NA	NA	20.2	6.98	02/03/2025	10:49
5	CCV1	1	Water	NA	NA	20.2	2.01	02/03/2025	10:50
6	Q1105-01	1	Solid	20.02	20	20.7	12.41	02/03/2025	11:00
7	Q1105-02	1	Solid	20.03	20	20.9	9.11	02/03/2025	11:10
8	Q1105-03	1	Solid	20.02	20	21.1	12.30	02/03/2025	11:15
9	Q1105-06	1	Solid	20.03	20	20.1	9.45	02/03/2025	11:25
10	Q1105-07	1	Solid	20.04	20	20.7	8.70	02/03/2025	11:30
11	Q1105-08	1	Solid	20.02	20	20.7	8.05	02/03/2025	11:35
12	Q1105-09	1	Solid	20.03	20	20.3	8.08	02/03/2025	11:40
13	Q1105-10	1	Solid	20.02	20	20.5	9.10	02/03/2025	11:45
14	Q1105-11	1	Solid	20.03	20	20.7	8.34	02/03/2025	11:50
15	Q1105-12	1	Solid	20.01	20	21.1	9.38	02/03/2025	12:00
16	CCV2	1	Water	NA	NA	20.2	12.02	02/03/2025	12:10
17	Q1105-13	1	Solid	20.04	20	20.9	8.55	02/03/2025	12:15
18	Q1105-14	1	Solid	20.02	20	20.7	8.65	02/03/2025	12:22
19	Q1105-15	1	Solid	20.03	20	20.7	8.42	02/03/2025	12:30
20	Q1105-16	1	Solid	20.04	20	20.3	8.41	02/03/2025	12:37

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	Q1105-17	1	Solid	20.02	20	20.6	8.45	02/03/2025	12:44
22	Q1105-18	1	Solid	20.03	20	20.8	7.00	02/03/2025	12:50
23	Q1105-19	1	Solid	20.02	20	21.2	7.92	02/03/2025	13:00
24	Q1105-20	1	Solid	20.03	20	21.0	7.82	02/03/2025	13:10
25	Q1105-21	1	Solid	20.02	20	21.2	8.62	02/03/2025	13:20
26	Q1105-22	1	Solid	20.03	20	20.5	8.41	02/03/2025	13:27
27	CCV3	1	Water	NA	NA	20.2	2.01	02/03/2025	13:35