

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134529

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3158

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	02/03/2025	09:10
2	Q1208-01		1	11	100.00	>212.0	>212.0	02/03/2025	09:40
3	Q1208-01DUP		1	11	100.00	>212.0	>212.0	02/03/2025	10:25
4	Q1271-03		1	12	100.00	>212.0	>212.0	02/03/2025	10:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06