

66134540

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

2/3/2025 16:48

Reviewed by : RM

Instrument ID : Konelab

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.988	0.0	0.121	
ICB1	-0.002	0.0	0.011	
CCV1	0.967	0.0	0.118	
CCB1	-0.003	0.0	0.011	
PB166347BL	-0.003	0.0	0.011	
PB166347BS	0.992	0.0	0.121	
Q1168-10	0.022	0.0	0.014	
Q1168-11	0.050	0.0	0.017	
CCV2	0.988	0.0	0.121	
CCB2	-0.002	0.0	0.011	

N	10
Mean	0.400
SD	0.5029
CV%	125.83

Aquakem v. 7.2AQ1

Results from time period:

Mon Feb 03 16:39:14 2025

Mon Feb 03 16:47:05 2025

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0028	mg/l	2/3/2025 15:09:51	
0.05PPM	A	Phenolics-	P	0.0517	mg/l	2/3/2025 15:09:52	
0.1PPM	A	Phenolics-	P	0.1008	mg/l	2/3/2025 15:09:53	
0.25PPM	A	Phenolics-	P	0.2566	mg/l	2/3/2025 15:09:54	
0.50PPM	A	Phenolics-	P	0.4998	mg/l	2/3/2025 15:09:55	
1.0PPM	A	Phenolics-	P	0.9895	mg/l	2/3/2025 15:09:56	
2.0PPM	A	Phenolics-	P	2.0044	mg/l	2/3/2025 15:09:57	
ICV1	S	Phenolics-	P	0.9878	mg/l	2/3/2025 16:39:14	
ICB1	S	Phenolics-	P	-0.0023	mg/l	2/3/2025 16:39:16	
CCV1	S	Phenolics-	P	0.9668	mg/l	2/3/2025 16:39:18	
CCB1	S	Phenolics-	P	-0.0031	mg/l	2/3/2025 16:39:20	
PB166347BL	S	Phenolics-	P	-0.0028	mg/l	2/3/2025 16:39:22	
PB166347BS	S	Phenolics-	P	0.9921	mg/l	2/3/2025 16:39:25	
Q1168-10	S	Phenolics-	P	0.0222	mg/l	2/3/2025 16:46:58	
Q1168-11	S	Phenolics-	P	0.0499	mg/l	2/3/2025 16:47:01	
CCV2	S	Phenolics-	P	0.9875	mg/l	2/3/2025 16:47:03	
CCB2	S	Phenolics-	P	-0.0017	mg/l	2/3/2025 16:47:04	

CHEMTECH CONSULTING GROUP INC
 284 Sheffield Street, Mountainside, NJ 07092

2/3/2025 15:16 Reviewed by : RM Instrument ID : Konelab


Test Phenolics-

Accepted 2/3/2025 15:16

Factor 9.052
 Bias 0.012

Coeff. of det. 0.999942

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.011	-0.0028	0.0000	-
2	Phenol-2	0.017	0.0517	0.0500	3.4
3	Phenol-2	0.023	0.1008	0.1000	0.8
4	Phenol-2	0.040	0.2566	0.2500	2.6
5	Phenol-2	0.067	0.4998	0.5000	0.0
6	Phenol-2	0.121	0.9895	1.0000	-1.1
7	Phenol-2	0.233	2.0044	2.0000	0.2

02/03/2025
 RM