

LB134542


7471B

INSTRUMENT ID: CY 1

Linear

 μ Abs.:

56180



A= 0.0000e+000

B= 1.7861e-004

C= -3.1927e-003

Rho= 0.9997269

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.058	0.058	340	0.000	340					1
0.2	0.200	0.195	-0.005	1111	0.0 %	1111					-3
2.5	2.500	2.380	-0.120	13343	0.0 %	13343					-5
5.0	5.000	5.130	0.130	28741	0.0 %	28741					3
7.5	7.500	7.406	-0.094	41479	0.0 %	41479					-1
10.0	10.000	10.031	0.031	56180	0.0 %	56180					0

LB134542 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Feb 2025 17:41:20

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	340	-	0.0000	7471B	PPB	03 Feb 2025 17:45:16	S	Std
0.2 - 1	50-2	1111	-	0.2000	7471B	PPB	03 Feb 2025 17:47:32	S	Std
2.5 - 1	52.5	13343	-	2.5000	7471B	PPB	03 Feb 2025 17:49:49	S	Std
5.0 - 1	55	28741	-	5.0000	7471B	PPB	03 Feb 2025 17:52:06	S	Std
7.5 - 1	57.5	41479	-	7.5000	7471B	PPB	03 Feb 2025 17:54:25	S	Std
10.0 - 1	510	56180	-	10.0000	7471B	PPB	03 Feb 2025 17:56:44	S	Std
ICV82 - 1	ICV82	21384	3.8163	-	7471B	PPB	03 Feb 2025 18:02:30	U	SMPL
ICB82 - 1	ICB82	106	0.0157	-	7471B	PPB	03 Feb 2025 18:04:45	U	SMPL
CCV14 - 1	CCV14	29019	5.1800	-	7471B	PPB	03 Feb 2025 18:07:02	U	SMPL
CCB14 - 1	CCB14	57	0.0070	-	7471B	PPB	03 Feb 2025 18:09:18	U	SMPL
CRA - 1	CRA	1083	0.1902	-	7471B	PPB	03 Feb 2025 18:11:36	U	SMPL
HighStd - 1	HighStd	55614	9.9302	-	7471B	PPB	03 Feb 2025 18:13:51	U	SMPL
ChkStd - 1	ChkStd	39134	6.9867	-	7471B	PPB	03 Feb 2025 18:16:07	U	SMPL
PB166520BL - 1	PBS	-90	-0.0193	-	7471B	PPB	03 Feb 2025 18:18:30	U	SMPL
PB166520BS - 1	LCSS	18792	3.3533	-	7471B	PPB	03 Feb 2025 18:20:53	U	SMPL
Q1215-03 - 1	JPP-29.1-012825	30165	5.3847	-	7471B	PPB	03 Feb 2025 18:23:09	U	SMPL
Q1215-07 - 1	JPP-29.2-012825	46571	8.3150	-	7471B	PPB	03 Feb 2025 18:25:28	U	SMPL
Q1241-03 - 1	JPP-3.5-013025	133169	23.7826	-	7471B	PPB	03 Feb 2025 18:27:47	U	SMPL
Q1241-07 - 1	JPP-5.3-013025	6738	1.2003	-	7471B	PPB	03 Feb 2025 18:30:08	U	SMPL
Q1241-11 - 1	JPP-5.2-013025	28766	5.1348	-	7471B	PPB	03 Feb 2025 18:32:42	U	SMPL
CCV15 - 1	CCV15	27962	4.9912	-	7471B	PPB	03 Feb 2025 18:35:01	U	SMPL
CCB15 - 1	CCB15	-206	-0.0400	-	7471B	PPB	03 Feb 2025 18:37:19	U	SMPL
Q1241-15 - 1	JPP-5.4-013025	73200	13.0713	-	7471B	PPB	03 Feb 2025 18:39:39	U	SMPL
Q1241-19 - 1	JPP-51.4-013025	43436	7.7551	-	7471B	PPB	03 Feb 2025 18:41:54	U	SMPL
Q1242-03 - 1	JPP-6.2-013025	36366	6.4923	-	7471B	PPB	03 Feb 2025 18:44:13	U	SMPL
Q1243-01 - 1	CL-01-01302025	576	0.0997	-	7471B	PPB	03 Feb 2025 18:46:32	U	SMPL
Q1244-01 - 1	EO-02-01302025	1588	0.2804	-	7471B	PPB	03 Feb 2025 18:48:51	U	SMPL
Q1254-01 - 1	OK-02-01312025	3018	0.5359	-	7471B	PPB	03 Feb 2025 18:51:07	U	SMPL
Q1261-01 - 1	CHRT-20430	842	0.1472	-	7471B	PPB	03 Feb 2025 18:53:23	U	SMPL
Q1262-01 - 1	ETGI-371	1458	0.2572	-	7471B	PPB	03 Feb 2025 18:55:41	U	SMPL
Q1262-03 - 1	CONCRETE-PAD	447	0.0766	-	7471B	PPB	03 Feb 2025 18:57:58	U	SMPL
Q1269-01 - 1	VNJ-231	1583	0.2796	-	7471B	PPB	03 Feb 2025 19:00:13	U	SMPL
CCV16 - 1	CCV16	29077	5.1904	-	7471B	PPB	03 Feb 2025 19:02:29	U	SMPL
CCB16 - 1	CCB16	-21	-0.0069	-	7471B	PPB	03 Feb 2025 19:04:44	U	SMPL
Q1269-01DUP - 1	VNJ-231DUP	1845	0.3263	-	7471B	PPB	03 Feb 2025 19:07:02	U	SMPL
Q1269-01MS - 1	VNJ-231MS	23444	4.1842	-	7471B	PPB	03 Feb 2025 19:09:17	U	SMPL
Q1269-01MSD - 1	VNJ-231MSD	20706	3.6952	-	7471B	PPB	03 Feb 2025 19:11:33	U	SMPL
Q1271-01 - 1	RBR200030	3120	0.5541	-	7471B	PPB	03 Feb 2025 19:13:52	U	SMPL
Q1271-03 - 1	3189-3196	126	0.0193	-	7471B	PPB	03 Feb 2025 19:16:09	U	SMPL
Q1269-01LX5 - 1		415	0.0709	-	7471B	PPB	03 Feb 2025 19:18:26	U	SMPL
Q1269-01A - 1		19823	3.5375	-	7471B	PPB	03 Feb 2025 19:20:42	U	SMPL
Q1241-03DLX5 - 1		25493	4.5502	-	7471B	PPB	03 Feb 2025 19:22:59	U	SMPL
Q1241-15DLX5 - 1		14514	2.5892	-	7471B	PPB	03 Feb 2025 19:25:19	U	SMPL
CCV17 - 1	CCV17	28062	5.0091	-	7471B	PPB	03 Feb 2025 19:27:37	U	SMPL
CCB17 - 1	CCB17	38	0.0036	-	7471B	PPB	03 Feb 2025 19:29:55	U	SMPL