

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134553

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3158

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	02/04/2025	13:15
2	Q1266-01		1	14	100.00	>212.0	>212.0	02/04/2025	14:00
3	Q1266-01DUP		1	14	100.00	>212.0	>212.0	02/04/2025	14:30
4	Q1266-02		1	11	100.00	>212.0	>212.0	02/04/2025	15:00
5	Q1266-03		1	13	100.00	>212.0	>212.0	02/04/2025	15:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06