

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134573

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3158

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	02/05/2025	08:20
2	Q1271-04		1	12	100.00	>212.0	>212.0	02/05/2025	08:50
3	Q1271-04DUP		1	12	100.00	>212.0	>212.0	02/05/2025	09:20
4	Q1282-01		1	14	100.00	>212.0	>212.0	02/05/2025	09:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6134573

WorkList Name : fp-2-5 WorkList ID : 187486 Department : Wet-Chemistry Date : 02-05-2025 08:07:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1271-04	3190	Water	Flash Point	Cool 4 deg C	PSEG03	D21	01/31/2025	1010B
Q1282-01	SW-WTS-02	Water	Flash Point	Cool 4 deg C	ENTA05	D11	01/31/2025	1010B

Date/Time 02/05/2025 08:10
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: RM CWL

Date/Time 02/05/2025 10:30
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: RM CWL