

LB134574


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

4835E



A= 0.0000e+000

B= 2.0257e-004 slope

C= -2.9773e-002 y-intercept

Rho= 0.9991109

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.00	0.000	-0.021	-0.021	44	0.000	44					0.61
0.05	0.050										1
0.20	0.200	0.194	-0.006	1103	0.0 %	1103					3
2.50	2.500	2.435	-0.065	12166	0.0 %	12166					3
5.00	5.000	5.036	0.036	25007	0.0 %	25007					1
7.50	7.500	7.791	0.291	38606	0.0 %	38606					4
10.0	10.000	9.766	-0.234	48358	0.0 %	48358					2

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	44	-		0 SFAM01.1	PPB	2/5/2025 10:55	Std
	0.2 S01	1103	-		0.2 SFAM01.1	PPB	2/5/2025 10:58	Std
	2.5 S02	12166	-		2.5 SFAM01.1	PPB	2/5/2025 11:00	Std
	5 S03	25007	-		5 SFAM01.1	PPB	2/5/2025 11:02	Std
	7.5 S04	38606	-		7.5 SFAM01.1	PPB	2/5/2025 11:04	Std
	10 S05	48358	-		10 SFAM01.1	PPB	2/5/2025 11:10	Std
ICV084	ICV084	19742	3.9693 -		SFAM01.1	PPB	2/5/2025 11:15	SMPL
ICB084	ICB084	-52	-0.0403 -		SFAM01.1	PPB	2/5/2025 11:17	SMPL
CCV021	CCV021	24795	4.9929 -		SFAM01.1	PPB	2/5/2025 11:19	SMPL
CCB021	CCB021	-163	-0.0628 -		SFAM01.1	PPB	2/5/2025 11:22	SMPL
PB166578BL	PBS578	69	-0.0158 -		SFAM01.1	PPB	2/5/2025 11:24	SMPL
Q1105-01	YE8G3	42	-0.0213 -		SFAM01.1	PPB	2/5/2025 11:26	SMPL
Q1105-02	YE8G5	2119	0.3995 -		SFAM01.1	PPB	2/5/2025 11:28	SMPL
Q1105-03	YE8G6	731	0.1183 -		SFAM01.1	PPB	2/5/2025 11:31	SMPL
Q1105-04	YE8G6D	692	0.1104 -		SFAM01.1	PPB	2/5/2025 11:33	SMPL
Q1105-05	YE8G6S	10818	2.1616 -		SFAM01.1	PPB	2/5/2025 11:35	SMPL
Q1105-06	YE8H0	367	0.0446 -		SFAM01.1	PPB	2/5/2025 11:38	SMPL
Q1105-07	YE8H1	5029	0.9889 -		SFAM01.1	PPB	2/5/2025 11:40	SMPL
Q1105-08	YE8H2	37899	7.6474 -		SFAM01.1	PPB	2/5/2025 11:42	SMPL
Q1105-09	YE8H3	25537	5.1432 -		SFAM01.1	PPB	2/5/2025 11:44	SMPL
Q1105-10	YE8H4	42	-0.0213 -		SFAM01.1	PPB	2/5/2025 11:47	SMPL
Q1105-11	YE8H5	4343	0.85 -		SFAM01.1	PPB	2/5/2025 11:49	SMPL
Q1105-12	YE8H6	3816	0.7432 -		SFAM01.1	PPB	2/5/2025 11:51	SMPL
Q1105-13	YE8E9	40966	8.2687 -		SFAM01.1	PPB	2/5/2025 11:53	SMPL
Q1105-14	YE8F0	99541	20.1341 -		SFAM01.1	PPB	2/5/2025 11:56	SMPL
Q1105-15	YE8F1	14257	2.8583 -		SFAM01.1	PPB	2/5/2025 11:58	SMPL
Q1105-16	YE8F2	54414	10.9928 -		SFAM01.1	PPB	2/5/2025 12:00	SMPL
Q1105-17	YE8F3	84553	17.098 -		SFAM01.1	PPB	2/5/2025 12:03	SMPL
Q1105-18	YE8F4	10001	1.9961 -		SFAM01.1	PPB	2/5/2025 12:05	SMPL
Q1105-19	YE8F6	2177	0.4112 -		SFAM01.1	PPB	2/5/2025 12:07	SMPL
CCV022	CCV022	23754	4.782 -		SFAM01.1	PPB	2/5/2025 12:10	SMPL
CCB022	CCB022	-124	-0.0549 -		SFAM01.1	PPB	2/5/2025 12:12	SMPL
Q1105-20	YE8G0	2763	0.5299 -		SFAM01.1	PPB	2/5/2025 12:14	SMPL
Q1105-21	YE8G1	1132	0.1995 -		SFAM01.1	PPB	2/5/2025 12:16	SMPL
Q1105-22	YE8G2	1466	0.2672 -		SFAM01.1	PPB	2/5/2025 12:19	SMPL
Q1105-14DLX5	YE8F0	19463	3.9128 -		SFAM01.1	PPB	2/5/2025 12:21	SMPL
Q1105-16DLX2	YE8F2	26735	5.3859 -		SFAM01.1	PPB	2/5/2025 12:23	SMPL
CCV023	CCV023	23882	4.808 -		SFAM01.1	PPB	2/5/2025 12:28	SMPL
CCB023	CCB023	-163	-0.0628 -		SFAM01.1	PPB	2/5/2025 12:30	SMPL
Q1105-17DLX5	YE8F3	15298	3.0691 -		SFAM01.1	PPB	2/5/2025 12:52	SMPL
CCV024	CCV024	23713	4.7737 -		SFAM01.1	PPB	2/5/2025 12:54	SMPL
CCB024	CCB024	-102	-0.0504 -		SFAM01.1	PPB	2/5/2025 12:57	SMPL