

LB134581


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35766



A= 0.0000e+000

B= 2.9120e-004

C= -1.1220e-001

Rho= 0.9957830

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.049	-0.049	217	0.000	217	0				
0.2	0.200	0.162	-0.038	942	0.0 %	942					-19
2.5	2.500	2.891	0.391	10312	0.0 %	10312					16
5.0	5.000	5.054	0.054	17742	0.0 %	17742					1
7.5	7.500	6.839	-0.661	23872	0.0 %	23872					-9
10.0	10.000	10.303	0.303	35766	0.0 %	35766					3

LB134581 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Feb 2025 14:01:15

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	217	-	0.0000	7470A	PPB	05 Feb 2025 14:35:37	S	Std
0.2 - 1	51.2	942	-	0.2000	7470A	PPB	05 Feb 2025 14:37:54	S	Std
2.5 - 1	52.5	10312	-	2.5000	7470A	PPB	05 Feb 2025 14:40:11	S	Std
5.0 - 1	55	17742	-	5.0000	7470A	PPB	05 Feb 2025 14:42:28	S	Std
7.5 - 1	52.5	23872	-	7.5000	7470A	PPB	05 Feb 2025 14:44:45	S	Std
10.0 - 1	516	35766	-	10.0000	7470A	PPB	05 Feb 2025 14:59:02	S	Std
ICV86 - 1	ICV86	15354	4.3588	-7470A	PPB	05 Feb 2025 15:02:01	U	SMPL	
ICB86 - 1	ICB86	64	-0.0936	-7470A	PPB	05 Feb 2025 15:04:16	U	SMPL	
CCV25 - 1	CCV25	16896	4.8079	-7470A	PPB	05 Feb 2025 15:12:43	U	SMPL	
CCB25 - 1	CCB25	16	-0.1075	-7470A	PPB	05 Feb 2025 15:14:58	U	SMPL	
CRA - 1	CRA	1182	0.2320	-7470A	PPB	05 Feb 2025 15:17:15	U	SMPL	
HighStd - 1	HighStd	35834	10.3226	-7470A	PPB	05 Feb 2025 15:19:31	U	SMPL	
ChkStd - 1	ChkStd	21591	6.1750	-7470A	PPB	05 Feb 2025 15:24:33	U	SMPL	
PB166572BL - 1	PBW	-89	-0.1381	-7470A	PPB	05 Feb 2025 15:26:48	U	SMPL	
PB166572BS - 1	LCSW	15691	4.4570	-7470A	PPB	05 Feb 2025 15:29:06	U	SMPL	
Q1279-01 - 1	COMP-1	-85	-0.1370	-7470A	PPB	05 Feb 2025 15:31:22	U	SMPL	
Q1279-02 - 1	COMP-2	266	-0.0347	-7470A	PPB	05 Feb 2025 15:33:40	U	SMPL	
Q1280-02 - 1	72-11984	473	0.0255	-7470A	PPB	05 Feb 2025 15:35:57	U	SMPL	
Q1285-02 - 1	72-11977	97	-0.0840	-7470A	PPB	05 Feb 2025 15:38:15	U	SMPL	
Q1287-02 - 1	BAY-DRUMS-A	-18	-0.1174	-7470A	PPB	05 Feb 2025 15:40:30	U	SMPL	
CCV26 - 1	CCV26	18876	5.3844	-7470A	PPB	05 Feb 2025 15:42:45	U	SMPL	
CCB26 - 1	CCB26	59	-0.0950	-7470A	PPB	05 Feb 2025 15:45:04	U	SMPL	
Q1287-04 - 1	BAY-DRUMS-B	17	-0.1072	-7470A	PPB	05 Feb 2025 15:47:21	U	SMPL	
Q1287-06 - 1	BAY-DRUMS-C	62	-0.0941	-7470A	PPB	05 Feb 2025 15:49:36	U	SMPL	
Q1287-08 - 1	BAY-DRUMS-D	90	-0.0860	-7470A	PPB	05 Feb 2025 15:51:52	U	SMPL	
Q1287-10 - 1	BAY-DRUMS-E	-7	-0.1142	-7470A	PPB	05 Feb 2025 15:54:08	U	SMPL	
Q1289-03 - 1	FL-DRUMS-A	-14	-0.1163	-7470A	PPB	05 Feb 2025 15:56:24	U	SMPL	
Q1289-05 - 1	FL-DRUMS-B	-4	-0.1134	-7470A	PPB	05 Feb 2025 15:58:40	U	SMPL	
Q1289-05DUP - 1	FL-DRUMS-BDUP	-35	-0.1224	-7470A	PPB	05 Feb 2025 16:00:57	U	SMPL	
Q1289-05MS - 1	FL-DRUMS-BMS	15071	4.2764	-7470A	PPB	05 Feb 2025 16:06:07	U	SMPL	
Q1289-05MSD - 1	FL-DRUMS-BMSD	15254	4.3297	-7470A	PPB	05 Feb 2025 16:08:24	U	SMPL	
CCV27 - 1	CCV27	18569	5.2950	-7470A	PPB	05 Feb 2025 16:20:24	U	SMPL	
CCB27 - 1	CCB27	-67	-0.1317	-7470A	PPB	05 Feb 2025 16:22:40	U	SMPL	
PB166528TB - 1	PB166528TB	49	-0.0979	-7470A	PPB	05 Feb 2025 16:25:00	U	SMPL	
Q1289-05LX5 - 1		88	-0.0866	-7470A	PPB	05 Feb 2025 16:27:15	U	SMPL	
Q1289-05A - 1		16446	4.6768	-7470A	PPB	05 Feb 2025 16:29:31	U	SMPL	
CCV28 - 1	CCV28	17773	5.0633	-7470A	PPB	05 Feb 2025 16:31:46	U	SMPL	
CCV28 - 1	CCV28	261	-0.0362	-7470A	PPB	05 Feb 2025 16:34:03	U	SMPL	