

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB134584

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/05/2025	15:00
2	CAL2	1	Water	NA	NA	20.3	7.01	02/05/2025	15:05
3	CAL3	1	Water	NA	NA	20.3	10.02	02/05/2025	15:10
4	ICV	1	Water	NA	NA	20.1	6.98	02/05/2025	15:15
5	CCV1	1	Water	NA	NA	20.2	2.01	02/05/2025	15:20
6	Q1282-01	1	Water	NA	NA	20.1	5.80	02/05/2025	15:25
7	Q1293-01	1	Water	NA	NA	20.3	6.75	02/05/2025	15:33
8	Q1298-01	1	Water	NA	NA	20.2	7.05	02/05/2025	15:37
9	Q1298-02	1	Water	NA	NA	20.2	7.11	02/05/2025	15:40
10	Q1298-03	1	Water	NA	NA	20.4	7.06	02/05/2025	15:47
11	Q1298-05	1	Water	NA	NA	20.7	7.11	02/05/2025	15:55
12	Q1298-06	1	Water	NA	NA	20.9	7.12	02/05/2025	16:00
13	Q1298-07	1	Water	NA	NA	20.5	6.99	02/05/2025	16:10
14	Q1298-07DUP	1	Water	NA	NA	20.4	7.00	02/05/2025	16:11
15	CCV2	1	Water	NA	NA	20.3	12.02	02/05/2025	16:15

WORKLIST(Hardcopy Internal Chain)

134584

WorkList Name : PH W Q1282

WorkList ID : 187507

Department : Wet-Chemistry

Date : 02-05-2025 14:39:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1282-01	D SW-WTS-02	Water	pH	Cool 4 deg C	ENTA05	D11	01/31/2025	9040C
Q1293-01	NWB-2123	Water	pH	Cool 4 deg C	PSEG03	D11	02/05/2025	9040C
Q1298-01	0114-A	Water	pH	Cool 4 deg C	PSEG03	N41	02/05/2025	9040C
Q1298-02	0114-B	Water	pH	Cool 4 deg C	PSEG03	N41	02/05/2025	9040C
Q1298-03	0114-C	Water	pH	Cool 4 deg C	PSEG03	N41	02/05/2025	9040C
Q1298-05	0131-A	Water	pH	Cool 4 deg C	PSEG03	N41	02/05/2025	9040C
Q1298-06	0131-B	Water	pH	Cool 4 deg C	PSEG03	N41	02/05/2025	9040C
Q1298-07	0131-C	Water	pH	Cool 4 deg C	PSEG03	N41	02/05/2025	9040C

Date/Time 02/05/25 14:50

Raw Sample Received by: SS W20

Raw Sample Relinquished by: JT (SM)

Date/Time 02/05/25 14:40

Raw Sample Received by: JT (SM)

Raw Sample Relinquished by: SS W20