


LB134599

7470A

INSTRUMENT ID: CV1

 $\mu$  Abs.:

43845



Linear

A= 0.0000e+000

B= 2.2698e-004

C= -1.4852e-002

Rho= 0.9997696

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D  |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|-----|
| 0.0    | 0.000  | 0.042 | 0.042  | 252   | 0.000      | 252   | 0     |       |       |       | 0/0 |
| 0.2    | 0.200  | 0.208 | 0.008  | 982   | 0.0 %      | 982   |       |       |       |       | 4   |
| 2.5    | 2.500  | 2.439 | -0.061 | 10810 | 0.0 %      | 10810 |       |       |       |       | -2  |
| 5.0    | 5.000  | 4.924 | -0.076 | 21759 | 0.0 %      | 21759 |       |       |       |       | -2  |
| 7.5    | 7.500  | 7.649 | 0.149  | 33766 | 0.0 %      | 33766 |       |       |       |       | 2   |
| 10.0   | 10.000 | 9.937 | -0.063 | 43845 | 0.0 %      | 43845 |       |       |       |       | -1  |

## LB134599 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Feb 2025 10:54:39

| Sample ID       | Extended ID                 | $\mu$ Abs | Conc    | Std Conc Method | Units | Date                 | Type | Type |
|-----------------|-----------------------------|-----------|---------|-----------------|-------|----------------------|------|------|
| 0.0 - 1         | 30                          | 252       | -       | 0.00007470A     | PPB   | 06 Feb 2025 11:22:56 | S    | Std  |
| 0.2 - 1         | 50.2                        | 982       | -       | 0.20007470A     | PPB   | 06 Feb 2025 11:25:13 | S    | Std  |
| 2.5 - 1         | 52.5                        | 10810     | -       | 2.50007470A     | PPB   | 06 Feb 2025 11:30:08 | S    | Std  |
| 5.0 - 1         | 55                          | 21759     | -       | 5.00007470A     | PPB   | 06 Feb 2025 11:34:59 | S    | Std  |
| 7.5 - 1         | 57.5                        | 33766     | -       | 7.50007470A     | PPB   | 06 Feb 2025 11:39:52 | S    | Std  |
| 10.0 - 1        | 59                          | 43845     | -       | 10.00007470A    | PPB   | 06 Feb 2025 11:47:28 | S    | Std  |
| ICV88 - 1       | ICV88                       | 16813     | 3.8014  | -7470A          | PPB   | 06 Feb 2025 11:52:45 | U    | SMPL |
| ICB88 - 1       | ICB88                       | -203      | -0.0609 | -7470A          | PPB   | 06 Feb 2025 11:55:01 | U    | SMPL |
| CCV30 - 1       | CCV30                       | 21670     | 4.9036  | -7470A          | PPB   | 06 Feb 2025 11:57:18 | U    | SMPL |
| CCB30 - 1       | CCB30                       | -156      | -0.0503 | -7470A          | PPB   | 06 Feb 2025 11:59:33 | U    | SMPL |
| CRA - 1         | CRA                         | 938       | 0.1981  | -7470A          | PPB   | 06 Feb 2025 12:01:51 | U    | SMPL |
| HighStd - 1     | HighStd                     | 44211     | 10.0203 | -7470A          | PPB   | 06 Feb 2025 12:04:05 | U    | SMPL |
| ChkStd - 1      | ChkStd                      | 31997     | 7.2479  | -7470A          | PPB   | 06 Feb 2025 12:08:42 | U    | SMPL |
| Q1201-01 - 1    | TAPHHA-MW12-012725-00-T2    | 294       | 0.0519  | -7470A          | PPB   | 06 Feb 2025 12:10:58 | U    | SMPL |
| Q1211-01 - 1    | TAPHHA-MW01-012825-00-T4    | 402       | 0.0764  | -7470A          | PPB   | 06 Feb 2025 12:13:15 | U    | SMPL |
| Q1211-02 - 1    | TAPIAL2-MW03-012825-00-T3   | 460       | 0.0896  | -7470A          | PPB   | 06 Feb 2025 12:15:32 | U    | SMPL |
| Q1266-01 - 1    | TRE-25-0004                 | 127       | 0.0140  | -7470A          | PPB   | 06 Feb 2025 12:17:48 | U    | SMPL |
| Q1266-02 - 1    | TRE-25-0010                 | 453       | 0.0880  | -7470A          | PPB   | 06 Feb 2025 12:20:05 | U    | SMPL |
| Q1266-03 - 1    | TRE-25-0012                 | 3         | -0.0142 | -7470A          | PPB   | 06 Feb 2025 12:22:22 | U    | SMPL |
| Q1271-04 - 1    | 3190                        | 423       | 0.0812  | -7470A          | PPB   | 06 Feb 2025 12:26:06 | U    | SMPL |
| CCV31 - 1       | CCV31                       | 20694     | 4.6823  | -7470A          | PPB   | 06 Feb 2025 12:28:21 | U    | SMPL |
| CCB31 - 1       | CCB31                       | -143      | -0.0473 | -7470A          | PPB   | 06 Feb 2025 12:30:36 | U    | SMPL |
| Q1282-01 - 1    | SW-WTS-02                   | 99        | 0.0076  | -7470A          | PPB   | 06 Feb 2025 12:32:53 | U    | SMPL |
| Q1201-01DUP - 1 | TAPHHA-MW12-012725-00-T2DUP | 135       | 0.0158  | -7470A          | PPB   | 06 Feb 2025 12:35:09 | U    | SMPL |
| Q1201-01MS - 1  | TAPHHA-MW12-012725-00-T2MS  | 14508     | 3.2782  | -7470A          | PPB   | 06 Feb 2025 12:37:24 | U    | SMPL |
| Q1201-01MSD - 1 | TAPHHA-MW12-012725-00-T2MSD | 13839     | 3.1264  | -7470A          | PPB   | 06 Feb 2025 12:39:40 | U    | SMPL |
| Q1201-01LX5 - 1 |                             | -46       | -0.0253 | -7470A          | PPB   | 06 Feb 2025 12:41:57 | U    | SMPL |
| Q1201-01A - 1   |                             | 15951     | 3.6058  | -7470A          | PPB   | 06 Feb 2025 12:44:15 | U    | SMPL |
| PB166596BL - 1  | PBW                         | 87        | 0.0049  | -7470A          | PPB   | 06 Feb 2025 12:49:04 | U    | SMPL |
| PB166596BS - 1  | LCSW                        | 16840     | 3.8075  | -7470A          | PPB   | 06 Feb 2025 12:56:43 | U    | SMPL |
| CCV32 - 1       | CCV32                       | 22628     | 5.1213  | -7470A          | PPB   | 06 Feb 2025 12:59:01 | U    | SMPL |
| CCB32 - 1       | CCB32                       | -46       | -0.0253 | -7470A          | PPB   | 06 Feb 2025 13:01:18 | U    | SMPL |