

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB134602

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/06/2025	13:25
2	CAL2	1	Water	NA	NA	20.2	7.00	02/06/2025	13:26
3	CAL3	1	Water	NA	NA	20.2	10.02	02/06/2025	13:30
4	ICV	1	Water	NA	NA	20.3	7.01	02/06/2025	13:32
5	CCV1	1	Water	NA	NA	20.2	2.01	02/06/2025	13:37
6	Q1320-01	1	Water	NA	NA	20.3	7.05	02/06/2025	13:45
7	Q1320-01DUP	1	Water	NA	NA	20.4	7.06	02/06/2025	13:46
8	Q1320-02	1	Water	NA	NA	20.1	7.75	02/06/2025	13:55
9	Q1320-03	1	Water	NA	NA	20.2	7.83	02/06/2025	14:10
10	CCV2	1	Water	NA	NA	20.2	12.02	02/06/2025	14:13

WORKLIST(Hardcopy Internal Chain)

VB 134602

WorkList Name : ph water q1320

WorkList ID : 187535

Department : Wet-Chemistry

Date : 02-06-2025 13:03:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1320-01 A	11725	Water	pH	Cool 4 deg C	PSEG03	N41	02/06/2025	9040C
Q1320-02 A	2325-A	Water	pH	Cool 4 deg C	PSEG03	N41	02/06/2025	9040C
Q1320-03 A	2325-B	Water	pH	Cool 4 deg C	PSEG03	N41	02/06/2025	9040C

Date/Time 02/06/25 13:15

Raw Sample Received by: RB 66001

Raw Sample Relinquished by: CP SM

Date/Time 02/06/25 15:00

Raw Sample Received by: CP SM

Raw Sample Relinquished by: RB 66001