

L8134603


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

46709



A= 0.0000e+000

B= 2.1840e-004

C= -6.9019e-002

Rho= 0.9994864

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	C/D
0.0	0.000	0.000	0.000	318	0.000	318					1
0.2	0.200	0.189	-0.011	1181	0.0 %	1181					6
2.5	2.500	2.661	0.161	12499	0.0 %	12499					6
5.0	5.000	4.875	-0.125	22639	0.0 %	22639					3
7.5	7.500	7.342	-0.158	33934	0.0 %	33934					2
10.0	10.000	10.132	0.132	46709	0.0 %	46709					1

LB134603 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Feb 2025 13:14:05

Sample ID	Extended ID	µ Abs	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	318	-	0.0000	7471B	PPB	06 Feb 2025 13:20:49	S	Std
0.2 - 1	SP	1181	-	0.2000	7471B	PPB	06 Feb 2025 13:25:42	S	Std
2.5 - 1	SP	12499	-	2.5000	7471B	PPB	06 Feb 2025 13:27:59	S	Std
5.0 - 1	SP	22639	-	5.0000	7471B	PPB	06 Feb 2025 13:30:16	S	Std
7.5 - 1	SP	33934	-	7.5000	7471B	PPB	06 Feb 2025 13:32:35	S	Std
10.0 - 1	SP	46709	-	10.0000	7471B	PPB	06 Feb 2025 13:37:42	S	Std
ICV89 - 1	ICV89	17164	3.6796	-	7471B	PPB	06 Feb 2025 13:40:33	U	SMPL
ICB89 - 1	ICB89	204	-0.0245	-	7471B	PPB	06 Feb 2025 13:42:49	U	SMPL
CCV33 - 1	CCV33	24614	5.3067	-	7471B	PPB	06 Feb 2025 13:45:07	U	SMPL
CCB33 - 1	CCB33	97	-0.0478	-	7471B	PPB	06 Feb 2025 13:47:22	U	SMPL
CRA - 1	CRA	1131	0.1780	-	7471B	PPB	06 Feb 2025 13:49:43	U	SMPL
HighStd - 1	HighStd	47776	10.3653	-	7471B	PPB	06 Feb 2025 13:52:00	U	SMPL
ChkStd - 1	ChkStd	32844	7.1042	-	7471B	PPB	06 Feb 2025 13:54:21	U	SMPL
PB166584BL - 1	PBS	327	0.0024	-	7471B	PPB	06 Feb 2025 14:03:35	U	SMPL
PB166584BS - 1	LCSS	18487	3.9686	-	7471B	PPB	06 Feb 2025 14:05:53	U	SMPL
Q1276-01 - 1	TR-05-020325	1235	0.2007	-	7471B	PPB	06 Feb 2025 14:08:09	U	SMPL
Q1277-02 - 1	RT 3249	3858	0.7738	-	7471B	PPB	06 Feb 2025 14:10:29	U	SMPL
Q1280-01 - 1	72-11984	4608	0.9374	-	7471B	PPB	06 Feb 2025 14:12:49	U	SMPL
Q1281-01 - 1	OR-02-2-03-2025	1951	0.3571	-	7471B	PPB	06 Feb 2025 14:15:09	U	SMPL
Q1284-01 - 1	SU-03-2-4-2025	1504	0.2595	-	7471B	PPB	06 Feb 2025 14:17:27	U	SMPL
CCV34 - 1	CCV34	24056	5.1848	-	7471B	PPB	06 Feb 2025 14:19:45	U	SMPL
CCB34 - 1	CCB34	605	0.0631	-	7471B	PPB	06 Feb 2025 14:22:00	U	SMPL
Q1285-01 - 1	72-11977	5084	1.0413	-	7471B	PPB	06 Feb 2025 14:24:18	U	SMPL
Q1286-01 - 1	NB-07-020425	830	0.1123	-	7471B	PPB	06 Feb 2025 14:26:33	U	SMPL
Q1287-01 - 1	BAY-DRUMS-A	2364	0.4473	-	7471B	PPB	06 Feb 2025 14:28:50	U	SMPL
Q1287-03 - 1	BAY-DRUMS-B	600	0.0620	-	7471B	PPB	06 Feb 2025 14:31:05	U	SMPL
Q1287-05 - 1	BAY-DRUMS-C	2520	0.4814	-	7471B	PPB	06 Feb 2025 14:33:22	U	SMPL
Q1287-07 - 1	BAY-DRUMS-D	1022	0.1542	-	7471B	PPB	06 Feb 2025 14:35:39	U	SMPL
Q1287-09 - 1	BAY-DRUMS-E	4645	0.9455	-	7471B	PPB	06 Feb 2025 14:37:56	U	SMPL
Q1288-01 - 1	HR-01-020425	5593	1.1525	-	7471B	PPB	06 Feb 2025 14:45:10	U	SMPL
Q1288-01DUP - 1	HR-01-020425DUP	5665	1.1682	-	7471B	PPB	06 Feb 2025 14:47:26	U	SMPL
Q1288-01MS - 1	HR-01-020425MS	26854	5.7959	-	7471B	PPB	06 Feb 2025 14:49:44	U	SMPL
CCV35 - 1	CCV35	24625	5.3091	-	7471B	PPB	06 Feb 2025 14:52:01	U	SMPL
CCB35 - 1	CCB35	516	0.0437	-	7471B	PPB	06 Feb 2025 14:54:22	U	SMPL
Q1288-01MSD - 1	HR-01-020425MSD	25245	5.4445	-	7471B	PPB	06 Feb 2025 14:56:41	U	SMPL
Q1288-04 - 1	HR-04-020425	28742	6.2083	-	7471B	PPB	06 Feb 2025 14:58:56	U	SMPL
Q1289-02 - 1	FL-DRUMS-A	-64	-0.0830	-	7471B	PPB	06 Feb 2025 15:01:17	U	SMPL
Q1289-04 - 1	FL-DRUMS-B	334	0.0039	-	7471B	PPB	06 Feb 2025 15:03:37	U	SMPL
Q1295-01 - 1	TAPFTA-SB01D-020425-00-T1	1178	0.1883	-	7471B	PPB	06 Feb 2025 15:05:53	U	SMPL
Q1304-01 - 1	RB21140	2727	0.5266	-	7471B	PPB	06 Feb 2025 15:08:09	U	SMPL
Q1288-01LX5 - 1		1320	0.2193	-	7471B	PPB	06 Feb 2025 15:10:26	U	SMPL
Q1288-01A - 1		24582	5.2997	-	7471B	PPB	06 Feb 2025 15:12:43	U	SMPL
CCV36 - 1	CCV36	24540	5.2906	-	7471B	PPB	06 Feb 2025 15:15:01	U	SMPL
CCB36 - 1	CCB36	461	0.0317	-	7471B	PPB	06 Feb 2025 15:17:20	U	SMPL