

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134605

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3158

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	02/06/2025	09:45
2	Q1293-01		1	13	100.00	>212.0	>212.0	02/06/2025	10:10
3	Q1298-04		1	13	100.00	>212.0	>212.0	02/06/2025	10:40
4	Q1298-08		1	14	100.00	>212.0	>212.0	02/06/2025	11:10
5	Q1307-02		1	11	100.00	>212.0	>212.0	02/06/2025	11:40
6	Q1307-03		1	14	100.00	>212.0	>212.0	02/06/2025	12:10
7	Q1307-03DUP		1	14	100.00	>212.0	>212.0	02/06/2025	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-2-6

WorkList ID : 187517

Department : Wet-Chemistry

Date : 02-06-2025 08:38:51

66134605

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1293-01	NWB-2123	Water	Flash Point	Cool 4 deg C	PSEG03	D11	02/05/2025	1010B
Q1298-04	0114-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	N41	02/05/2025	1010B
Q1298-08	0131-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	N41	02/05/2025	1010B
Q1307-02	431	Water	Flash Point	Cool 4 deg C	PSEG03	N31	02/05/2025	1010B
Q1307-03	433	Water	Flash Point	Cool 4 deg C	PSEG03	N31	02/05/2025	1010B

Date/Time 02/06/2025 09:36
 Raw Sample Received by: RM CWL
 Raw Sample Relinquished by: [Signature]

Date/Time 02/06/2025 13:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: RM CWL