

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB134608

Slope : 98.4

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/06/2025	09:30
2	CAL2	1	Water	NA	NA	20.2	7.00	02/06/2025	09:31
3	CAL3	1	Water	NA	NA	20.3	10.02	02/06/2025	09:33
4	ICV	1	Water	NA	NA	20.2	7.01	02/06/2025	09:37
5	CCV1	1	Water	NA	NA	20.2	2.01	02/06/2025	09:40
6	Q1299-04	1	Solid	20.02	20	20.9	8.05	02/06/2025	09:50
7	Q1299-04DUP	1	Solid	20.03	20	20.8	8.06	02/06/2025	09:51
8	Q1299-08	1	Solid	20.04	20	21.2	8.56	02/06/2025	09:59
9	Q1307-01	1	Solid	20.02	20	20.6	6.06	02/06/2025	10:10
10	Q1309-04	1	Solid	20.03	20	21.5	5.55	02/06/2025	10:15
11	Q1309-08	1	Solid	20.02	20	22.3	9.11	02/06/2025	10:25
12	Q1309-12	1	Solid	20.04	20	21.8	6.91	02/06/2025	10:35
13	Q1309-16	1	Solid	20.03	20	21.6	8.50	02/06/2025	10:40
14	Q1309-20	1	Solid	20.02	20	21.0	7.96	02/06/2025	10:48
15	Q1309-24	1	Solid	20.03	20	21.4	7.51	02/06/2025	11:00
16	CCV2	1	Water	NA	NA	20.3	12.02	02/06/2025	11:10

WORKLIST(Hardcopy Internal Chain)

VB 124608

WorkList Name : corrosivity q1309

WorkList ID : 187523

Department : Wet-Chemistry

Date : 02-06-2025 09:05:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1299-04	WC-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	02/04/2025	9045D
Q1299-08	WC-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	02/04/2025	9045D
Q1307-01	434	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	02/05/2025	9045D
Q1309-04	WC-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	02/05/2025	9045D
Q1309-08	WC-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	02/05/2025	9045D
Q1309-12	WC-4	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	02/05/2025	9045D
Q1309-16	WC-5	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	02/05/2025	9045D
Q1309-20	WC-6	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	02/05/2025	9045D
Q1309-24	WC-8	Solid	Corrosivity	Cool 4 deg C	PSEG03	N31	02/05/2025	9045D

Date/Time 02/06/25 09:12:0

Raw Sample Received by: SB WDCI

Raw Sample Relinquished by: JH-OSR

Date/Time 02/06/25 12:30

Raw Sample Received by: JH-OSR

Raw Sample Relinquished by: SB WDCI