

LB134624


7471B

INSTRUMENT ID: CVZ

Linear

 μ Abs.:

41899



A= 0.0000e+000

B= 2.3898e-004

C= -1.2113e-001

Rho= 0.9993022

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.0	0.000	-0.064	-0.064	239	0.000	239					0.0
0.2	0.200	0.196	-0.004	1326	0.0 %	1326					-2
2.5	2.500	2.662	0.162	11646	0.0 %	11646					6
5.0	5.000	4.818	-0.182	20667	0.0 %	20667					-4
7.5	7.500	7.697	0.197	32714	0.0 %	32714					3
10.0	10.000	9.892	-0.108	41899	0.0 %	41899					1

LB134624 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Feb 2025 14:03:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	239	-	0.0000	7471B	PPB	07 Feb 2025 14:05:51	S	Std
0.2 - 1	102	1326	-	0.2000	7471B	PPB	07 Feb 2025 14:08:08	S	Std
2.5 - 1	12.5	11646	-	2.5000	7471B	PPB	07 Feb 2025 14:10:25	S	Std
5.0 - 1	55	20667	-	5.0000	7471B	PPB	07 Feb 2025 14:12:44	S	Std
7.5 - 1	37.5	32714	-	7.5000	7471B	PPB	07 Feb 2025 14:17:48	S	Std
10.0 - 1	377	41899	-	10.0000	7471B	PPB	07 Feb 2025 14:22:46	S	Std
ICV91 - 1	ICV91	15104	3.4884	-	7471B	PPB	07 Feb 2025 14:25:35	U	SMPL
ICB91 - 1	ICB91	0	-0.1211	-	7471B	PPB	07 Feb 2025 14:27:53	U	SMPL
ICV91 - 1	ICV91	16604	3.8468	-	7471B	PPB	07 Feb 2025 14:28:07	U	SMPL
ICB91 - 1	ICB91	-37	-0.1300	-	7471B	PPB	07 Feb 2025 14:30:24	U	SMPL
CCV41 - 1	CCV41	21922	5.1177	-	7471B	PPB	07 Feb 2025 14:32:46	U	SMPL
CCB41 - 1	CCB41	-48	-0.1326	-	7471B	PPB	07 Feb 2025 14:35:01	U	SMPL
CRA - 1	CRA	1231	0.1731	-	7471B	PPB	07 Feb 2025 14:37:23	U	SMPL
HighStd - 1	HighStd	42116	9.9436	-	7471B	PPB	07 Feb 2025 14:39:38	U	SMPL
ChkStd - 1	ChkStd	31218	7.3392	-	7471B	PPB	07 Feb 2025 14:41:54	U	SMPL
PB166625BL - 1	PBS	297	-0.0502	-	7471B	PPB	07 Feb 2025 14:47:47	U	SMPL
PB166625BS - 1	LCSS	17843	4.1429	-	7471B	PPB	07 Feb 2025 14:52:40	U	SMPL
Q1299-01 - 1	WC-1	3732	0.7707	-	7471B	PPB	07 Feb 2025 15:00:01	U	SMPL
Q1299-01DUP - 1	WC-1DUP	3555	0.7284	-	7471B	PPB	07 Feb 2025 15:02:18	U	SMPL
Q1299-01MS - 1	WC-1MS	22747	5.3149	-	7471B	PPB	07 Feb 2025 15:04:37	U	SMPL
Q1299-01MSD - 1	WC-1MSD	22823	5.3330	-	7471B	PPB	07 Feb 2025 15:06:55	U	SMPL
Q1299-05 - 1	WC-2	1623	0.2667	-	7471B	PPB	07 Feb 2025 15:09:14	U	SMPL
CCV42 - 1	CCV42	21498	5.0164	-	7471B	PPB	07 Feb 2025 15:11:35	U	SMPL
CCB42 - 1	CCB42	189	-0.0760	-	7471B	PPB	07 Feb 2025 15:13:51	U	SMPL
Q1309-01 - 1	WC-2	1542	0.2474	-	7471B	PPB	07 Feb 2025 15:16:12	U	SMPL
Q1309-05 - 1	WC-3	1397	0.2127	-	7471B	PPB	07 Feb 2025 15:18:28	U	SMPL
Q1309-09 - 1	WC-4	3579	0.7342	-	7471B	PPB	07 Feb 2025 15:20:44	U	SMPL
Q1309-13 - 1	WC-5	488	-0.0045	-	7471B	PPB	07 Feb 2025 15:23:01	U	SMPL
Q1309-17 - 1	WC-6	2011	0.3595	-	7471B	PPB	07 Feb 2025 15:25:18	U	SMPL
Q1309-21 - 1	WC-8	1089	0.1391	-	7471B	PPB	07 Feb 2025 15:27:34	U	SMPL
Q1299-01LX5 - 1		677	0.0407	-	7471B	PPB	07 Feb 2025 15:29:52	U	SMPL
Q1299-01A - 1		22745	5.3144	-	7471B	PPB	07 Feb 2025 15:41:56	U	SMPL
CCV43 - 1	CCV43	22857	5.3412	-	7471B	PPB	07 Feb 2025 15:44:13	U	SMPL
CCB43 - 1	CCB43	-84	-0.1412	-	7471B	PPB	07 Feb 2025 15:46:33	U	SMPL