

LB134650


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

39432



A= 0.0000e+000

B= 2.5504e-004

C= 6.0795e-002

Rho= 0.9995035

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/O
0.0	0.000	0.034	0.034	-107	0.000	-107					
0.2	0.200	0.191	-0.009	511	0.0 %	511	0				-5
2.5	2.500	2.639	0.139	10107	0.0 %	10107					6
5.0	5.000	4.798	-0.202	18576	0.0 %	18576					-4
7.5	7.500	7.421	-0.079	28858	0.0 %	28858					-1
10.0	10.000	10.118	0.118	39432	0.0 %	39432					1

Method: 7470A

Operator: Admin

LB134650 INSTRUMENT ID : CV1

Date of Analysis: 10 Feb 2025 12:47:04

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-107	-	0.000	7470A	PPB	10 Feb 2025 12:51:38	S	Std
0.2 - 1	51	511	-	0.200	7470A	PPB	10 Feb 2025 12:53:54	S	Std
2.5 - 1	52	10107	-	2.500	7470A	PPB	10 Feb 2025 12:56:12	S	Std
5.0 - 1	53	18576	-	5.000	7470A	PPB	10 Feb 2025 12:58:28	S	Std
7.5 - 1	54	28858	-	7.500	7470A	PPB	10 Feb 2025 13:00:46	S	Std
10.0 - 1	55	39432	-	10.000	7470A	PPB	10 Feb 2025 13:03:04	S	Std
ICV93 - 1	ICV93	14479	3.7535	-7470A	PPB	10 Feb 2025 13:05:54	U	SMPL	
ICB93 - 1	ICB93	-231	0.0019	-7470A	PPB	10 Feb 2025 13:08:10	U	SMPL	
CCV47 - 1	CCV47	19174	4.9510	-7470A	PPB	10 Feb 2025 13:10:32	U	SMPL	
CCB47 - 1	CCB47	-274	-0.0091	-7470A	PPB	10 Feb 2025 13:12:48	U	SMPL	
CRA - 1	CRA	670	0.2317	-7470A	PPB	10 Feb 2025 13:17:45	U	SMPL	
HighStd - 1	HighStd	40683	10.4367	-7470A	PPB	10 Feb 2025 13:20:00	U	SMPL	
ChkStd - 1	ChkStd	27124	6.9786	-7470A	PPB	10 Feb 2025 13:22:15	U	SMPL	
PB166621BL - 1	PBW	-428	-0.0484	-7470A	PPB	10 Feb 2025 13:24:34	U	SMPL	
PB166621BS - 1	LCSW	17400	4.4985	-7470A	PPB	10 Feb 2025 13:26:51	U	SMPL	
Q1299-04 - 1	WC-1	-212	0.0067	-7470A	PPB	10 Feb 2025 13:29:07	U	SMPL	
Q1299-08 - 1	WC-2	1794	0.5183	-7470A	PPB	10 Feb 2025 13:31:26	U	SMPL	
Q1307-01 - 1	434	240	0.1220	-7470A	PPB	10 Feb 2025 13:33:43	U	SMPL	
Q1309-04 - 1	WC-2	286	0.1337	-7470A	PPB	10 Feb 2025 13:36:00	U	SMPL	
Q1309-08 - 1	WC-3	-8	0.0588	-7470A	PPB	10 Feb 2025 13:38:16	U	SMPL	
CCV48 - 1	CCV48	18726	4.8367	-7470A	PPB	10 Feb 2025 13:40:31	U	SMPL	
CCB48 - 1	CCB48	-216	0.0057	-7470A	PPB	10 Feb 2025 13:42:46	U	SMPL	
Q1309-12 - 1	WC-4	111	0.0891	-7470A	PPB	10 Feb 2025 13:45:04	U	SMPL	
Q1309-16 - 1	WC-5	411	0.1656	-7470A	PPB	10 Feb 2025 13:47:19	U	SMPL	
Q1309-20 - 1	WC-6	152	0.0996	-7470A	PPB	10 Feb 2025 13:49:35	U	SMPL	
Q1309-24 - 1	WC-8	188	0.1087	-7470A	PPB	10 Feb 2025 13:51:50	U	SMPL	
Q1309-24DUP - 1	WC-8DUP	434	0.1715	-7470A	PPB	10 Feb 2025 13:54:05	U	SMPL	
Q1309-24MS - 1	WC-8MS	15099	3.9117	-7470A	PPB	10 Feb 2025 13:56:21	U	SMPL	
Q1309-24MSD - 1	WC-8MSD	13874	3.5992	-7470A	PPB	10 Feb 2025 13:58:41	U	SMPL	
PB166658BL - 1	PBW	12	0.0639	-7470A	PPB	10 Feb 2025 14:03:40	U	SMPL	
PB166658BS - 1	LCSW	16194	4.1909	-7470A	PPB	10 Feb 2025 14:05:57	U	SMPL	
Q1307-02 - 1	431	-156	0.0210	-7470A	PPB	10 Feb 2025 14:08:13	U	SMPL	
CCV49 - 1	CCV49	20648	5.3269	-7470A	PPB	10 Feb 2025 14:13:46	U	SMPL	
CCB49 - 1	CCB49	-232	0.0016	-7470A	PPB	10 Feb 2025 14:16:02	U	SMPL	
Q1307-02DUP - 1	431DUP	-56	0.0465	-7470A	PPB	10 Feb 2025 14:18:19	U	SMPL	
Q1307-02MS - 1	431MS	12462	3.2391	-7470A	PPB	10 Feb 2025 14:22:58	U	SMPL	
Q1307-02MSD - 1	431MSD	13404	3.4794	-7470A	PPB	10 Feb 2025 14:25:15	U	SMPL	
Q1307-03 - 1	433	1088	0.3383	-7470A	PPB	10 Feb 2025 14:27:33	U	SMPL	
Q1320-01 - 1	11725	2042	0.5816	-7470A	PPB	10 Feb 2025 14:29:51	U	SMPL	
PB166591TB - 1		-98	0.0358	-7470A	PPB	10 Feb 2025 14:32:10	U	SMPL	
Q1309-24LX5 - 1		12	0.0639	-7470A	PPB	10 Feb 2025 14:34:37	U	SMPL	
Q1309-24A - 1		14321	3.7133	-7470A	PPB	10 Feb 2025 14:36:54	U	SMPL	
Q1307-02LX5 - 1		-182	0.0144	-7470A	PPB	10 Feb 2025 14:39:11	U	SMPL	
Q1307-02A - 1		14175	3.6760	-7470A	PPB	10 Feb 2025 14:41:29	U	SMPL	
CCV50 - 1	CCV50	20214	5.2162	-7470A	PPB	10 Feb 2025 14:46:22	U	SMPL	
CCB50 - 1	CCB50	-320	-0.0208	-7470A	PPB	10 Feb 2025 14:48:39	U	SMPL	