



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 2/11/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:45
In Date: 02/10/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:37
Out Date: 02/11/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB134651

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q1344-01	SOIL-1	1	1.15	8.81	9.96	9.1	90.2	
Q1344-02	SOIL-1	2	1.15	8.81	9.96	9.1	90.2	
Q1344-03	SOIL-2	3	1.16	8.81	9.97	9.00	89.0	
Q1344-04	SOIL-2	4	1.16	8.81	9.97	9.00	89.0	
Q1345-01	EO-01-021025	5	1.12	8.75	9.87	9.14	91.7	
Q1345-02	EO-01-021025-E2	6	1.13	8.54	9.67	8.82	90.0	
Q1346-01	SOIL-COMP	7	1.14	8.55	9.69	8.56	86.8	
Q1346-03	CONCRETE	8	1.00	1.00	2.00	2.00	100.0	concrete sample, 100% solids

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$