

LB134657


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46134



A= 0.0000e+000

B= 2.1544e-004 slope

C= -3.6487e-002 y-intercept

Rho= 0.9998355

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/d)
0.00	0.000	0.009	0.009	210	0.000	210					=
0.05	0.050										-13
0.20	0.200	0.174	-0.026	975	0.0 %	975					-1
2.50	2.500	2.474	-0.026	11655	0.0 %	11655					0
5.00	5.000	5.015	0.015	23449	0.0 %	23449					2
7.50	7.500	7.625	0.125	35564	0.0 %	35564					-1
10.0	10.000	9.903	-0.097	46134	0.0 %	46134					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0	210	-		0 SFAM01.1	PPB	2/10/2025 16:10	Std
	0.2	975	-		0.2 SFAM01.1	PPB	2/10/2025 16:12	Std
	2.5	11655	-		2.5 SFAM01.1	PPB	2/10/2025 16:14	Std
	5	23449	-		5 SFAM01.1	PPB	2/10/2025 16:17	Std
	7.5	35564	-		7.5 SFAM01.1	PPB	2/10/2025 16:19	Std
	10	46134	-		10 SFAM01.1	PPB	2/10/2025 16:24	Std
ICV094	ICV094	18103	3.8636 -		SFAM01.1	PPB	2/10/2025 16:29	SMPL
ICB094	ICB094	69	-0.0216 -		SFAM01.1	PPB	2/10/2025 16:31	SMPL
CCV051	CCV051	23868	5.1056 -		SFAM01.1	PPB	2/10/2025 16:34	SMPL
CCB051	CCB051	12	-0.0339 -		SFAM01.1	PPB	2/10/2025 16:36	SMPL
PB166633BL	PBS633	65	-0.0225 -		SFAM01.1	PPB	2/10/2025 16:38	SMPL
Q1151-01	MC0AZ0	31536	6.7576 -		SFAM01.1	PPB	2/10/2025 16:40	SMPL
Q1151-02	MC0AZ1	20941	4.475 -		SFAM01.1	PPB	2/10/2025 16:43	SMPL
Q1151-03	MC0AZ2	12432	2.6418 -		SFAM01.1	PPB	2/10/2025 16:45	SMPL
Q1151-04	MC0AZ3	2182	0.4336 -		SFAM01.1	PPB	2/10/2025 16:47	SMPL
Q1151-05	MC0AZ4	88873	19.1102 -		SFAM01.1	PPB	2/10/2025 16:50	SMPL
Q1151-06	MC0AZ5	9110	1.9262 -		SFAM01.1	PPB	2/10/2025 16:52	SMPL
Q1151-07	MC0AZ6	2651	0.5346 -		SFAM01.1	PPB	2/10/2025 16:54	SMPL
Q1151-08	MC0AZ7	3149	0.6419 -		SFAM01.1	PPB	2/10/2025 16:57	SMPL
Q1151-09	MC0AZ7D	2619	0.5277 -		SFAM01.1	PPB	2/10/2025 16:59	SMPL
Q1151-10	MC0AZ7S	13924	2.9633 -		SFAM01.1	PPB	2/10/2025 17:01	SMPL
Q1151-11	MC0B00	50458	10.8341 -		SFAM01.1	PPB	2/10/2025 17:04	SMPL
Q1151-12	MC0B01	3659	0.7518 -		SFAM01.1	PPB	2/10/2025 17:06	SMPL
Q1159-01	YE8C9	3086	0.6284 -		SFAM01.1	PPB	2/10/2025 17:08	SMPL
Q1159-02	YE8C9D	3437	0.704 -		SFAM01.1	PPB	2/10/2025 17:11	SMPL
Q1159-03	YE8C9S	15327	3.2655 -		SFAM01.1	PPB	2/10/2025 17:13	SMPL
Q1159-04	YE8D1	5829	1.2193 -		SFAM01.1	PPB	2/10/2025 17:15	SMPL
Q1159-05	YE8D3	10147	2.1496 -		SFAM01.1	PPB	2/10/2025 17:17	SMPL
Q1159-06	YE8E2	700	0.1143 -		SFAM01.1	PPB	2/10/2025 17:20	SMPL
CCV052	CCV052	24452	5.2314 -		SFAM01.1	PPB	2/10/2025 17:22	SMPL
CCB052	CCB052	-53	-0.0479 -		SFAM01.1	PPB	2/10/2025 17:24	SMPL
Q1151-05DLX5	MC0AZ4	18099	3.8627 -		SFAM01.1	PPB	2/10/2025 17:27	SMPL
Q1151-11DLX2	MC0B00	25850	5.5326 -		SFAM01.1	PPB	2/10/2025 17:29	SMPL
CCV053	CCV053	23383	5.0011 -		SFAM01.1	PPB	2/10/2025 17:31	SMPL
CCB053	CCB053	-53	-0.0479 -		SFAM01.1	PPB	2/10/2025 17:33	SMPL