

66134650

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

2/10/2025 16:47

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.007	0.0	0.102	
ICB1	-0.003	0.0	0.003	
CCV1	1.023	0.0	0.104	
CCB1	-0.004	0.0	0.003	
PB166645BL	-0.006	0.0	0.003	
PB166645BS	1.015	0.0	0.103	
Q1168-03	0.024	0.0	0.005	
PB166646BL	-0.001	0.0	0.003	
PB166646BS	1.013	0.0	0.103	
Q1168-09	0.023	0.0	0.005	
CCV2	1.016	0.0	0.103	
CCB2	-0.002	0.0	0.003	

N	12
Mean	0.426
SD	0.5205
CV%	122.31

Aquakem v. 7.2AQ1

Results from time period:

Mon Feb 10 16:29:34 2025

Mon Feb 10 16:43:35 2025

Sample Id	Sam/Ctr/c/	Test short r	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0042	mg/l	2/10/2025 15:17:27	
0.05PPM	A	Phenolics-	P	0.0485	mg/l	2/10/2025 15:17:28	
0.1PPM	A	Phenolics-	P	0.1025	mg/l	2/10/2025 15:17:29	
0.25PPM	A	Phenolics-	P	0.2508	mg/l	2/10/2025 15:17:30	
0.50PPM	A	Phenolics-	P	0.5049	mg/l	2/10/2025 15:17:31	
1.0PPM	A	Phenolics-	P	0.9979	mg/l	2/10/2025 15:17:32	
2.0PPM	A	Phenolics-	P	1.9996	mg/l	2/10/2025 15:17:33	
ICV1	S	Phenolics-	P	1.0073	mg/l	2/10/2025 16:29:34	
ICB1	S	Phenolics-	P	-0.0027	mg/l	2/10/2025 16:29:36	
CCV1	S	Phenolics-	P	1.0228	mg/l	2/10/2025 16:29:38	
CCB1	S	Phenolics-	P	-0.0042	mg/l	2/10/2025 16:29:40	
PB166645BL	S	Phenolics-	P	-0.0057	mg/l	2/10/2025 16:29:43	
PB166645BS	S	Phenolics-	P	1.0154	mg/l	2/10/2025 16:29:44	
Q1168-03	S	Phenolics-	P	0.024	mg/l	2/10/2025 16:39:36	
PB166646BL	S	Phenolics-	P	-0.001	mg/l	2/10/2025 16:39:38	
PB166646BS	S	Phenolics-	P	1.0133	mg/l	2/10/2025 16:39:41	
Q1168-09	S	Phenolics-	P	0.0232	mg/l	2/10/2025 16:39:43	
CCV2	S	Phenolics-	P	1.0161	mg/l	2/10/2025 16:39:44	
CCB2	S	Phenolics-	P	-0.0023	mg/l	2/10/2025 16:39:46	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

2/10/2025 15:19

Reviewed by : RM

Instrument ID : Konelab


Test Phenolics-

Accepted 2/10/2025 15:19

Factor 10.16
Bias 0.003

Coeff. of det. 0.999982

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.003	-0.0042	0.0000	-
2	Phenol-2	0.008	0.0485	0.0500	-3.0
3	Phenol-2	0.013	0.1025	0.1000	2.5
4	Phenol-2	0.028	0.2508	0.2500	0.3
5	Phenol-2	0.053	0.5049	0.5000	1.0
6	Phenol-2	0.101	0.9979	1.0000	-0.2
7	Phenol-2	0.200	1.9996	2.0000	0.0

02/10/2025
RM