

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134676

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/12/2025	08:25
2	CAL2	1	Water	NA	NA	20.2	7.01	02/12/2025	08:26
3	CAL3	1	Water	NA	NA	20.3	10.02	02/12/2025	08:30
4	ICV	1	Water	NA	NA	20.2	6.98	02/12/2025	08:33
5	CCV1	1	Water	NA	NA	20.3	2.02	02/12/2025	09:00
6	Q1348-01	1	Water	NA	NA	20.7	6.52	02/12/2025	09:10
7	Q1356-04	1	Water	NA	NA	20.9	7.06	02/12/2025	09:20
8	Q1356-05	1	Water	NA	NA	20.9	5.94	02/12/2025	09:25
9	Q1356-06	1	Water	NA	NA	20.8	6.73	02/12/2025	09:35
10	Q1356-07	1	Water	NA	NA	20.2	7.62	02/12/2025	09:40
11	Q1356-08	1	Water	NA	NA	20.2	6.61	02/12/2025	09:45
12	Q1356-09	1	Water	NA	NA	20.8	6.37	02/12/2025	09:50
13	Q1356-09DUP	1	Water	NA	NA	20.9	6.38	02/12/2025	09:51
14	CCV2	1	Water	NA	NA	20.3	12.02	02/12/2025	10:00

WORKLIST(Hardcopy Internal Chain)

13134676

WorkList Name : ph w q1356 WorkList ID : 187649 Department : Wet-Chemistry Date : 02-12-2025 08:03:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1348-01	TWP-1-WC	Water	pH	Cool 4 deg C	PSEG03	N51	02/10/2025	9040C
Q1356-04	CARBON-WATER	Water	pH	Cool 4 deg C	PSEG04	N51	02/11/2025	9040C
Q1356-05	CARBON-FB	Water	pH	Cool 4 deg C	PSEG04	N51	02/11/2025	9040C
Q1356-06	WATER-A	Water	pH	Cool 4 deg C	PSEG04	N51	02/11/2025	9040C
Q1356-07	WATER-B	Water	pH	Cool 4 deg C	PSEG04	N51	02/11/2025	9040C
Q1356-08	WATER-FB	Water	pH	Cool 4 deg C	PSEG04	N51	02/11/2025	9040C
Q1356-09	SOIL-FB	Water	pH	Cool 4 deg C	PSEG04	N51	02/11/2025	9040C

Date/Time 02/12/25 08:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 02/12/25 13:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]