

LB134678


SFAM01.1

INSTRUMENT ID: CYL

Linear

 μ Abs.:

36405



A= 0.0000e+000

B= 2.7698e-004 slope

C= -6.8860e-002 y-intercept

Rho= 0.9999588

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/d
0.00	0.000	0.004	0.004	262	0.000	262					
0.05	0.050										
0.20	0.200	0.177	-0.023	886	0.0 %	886					
2.50	2.500	2.512	0.012	9319	0.0 %	9319					
5.00	5.000	5.050	0.050	18481	0.0 %	18481					
7.50	7.500	7.443	-0.057	27119	0.0 %	27119					
10.0	10.000	10.015	0.015	36405	0.0 %	36405					

9/11
112
01-010

LB134678

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	262	-		0 SFAM01.1	PPB	2/12/2025 10:37	Std
	0.2 S01	886	-		0.2 SFAM01.1	PPB	2/12/2025 10:40	Std
	2.5 S02	9319	-		2.5 SFAM01.1	PPB	2/12/2025 10:44	Std
	5 S03	18481	-		5 SFAM01.1	PPB	2/12/2025 10:46	Std
	7.5 S04	27119	-		7.5 SFAM01.1	PPB	2/12/2025 10:48	Std
	10 S05	36405	-		10 SFAM01.1	PPB	2/12/2025 10:59	Std
ICV001	ICV001	14651	3.9892 -		SFAM01.1	PPB	2/12/2025 11:04	SMPL
ICB001	ICB001	254	0.0015 -		SFAM01.1	PPB	2/12/2025 11:09	SMPL
CCV061	CCV061	18673	5.1032 -		SFAM01.1	PPB	2/12/2025 11:11	SMPL
CCB061	CCB061	279	0.0084 -		SFAM01.1	PPB	2/12/2025 11:27	SMPL
PBW1	PBW1	324	0.0209 -		SFAM01.1	PPB	2/12/2025 11:29	SMPL
PBW2	PBW2	261	0.0034 -		SFAM01.1	PPB	2/12/2025 11:34	SMPL
PBW3	PBW3	259	0.0029 -		SFAM01.1	PPB	2/12/2025 11:36	SMPL
MDL1	MDL1	391	0.0394 -		SFAM01.1	PPB	2/12/2025 11:41	SMPL
MDL2	MDL2	456	0.0574 -		SFAM01.1	PPB	2/12/2025 11:43	SMPL
MDL3	MDL3	468	0.0608 -		SFAM01.1	PPB	2/12/2025 11:45	SMPL
CCV062	CCV062	19336	5.2869 -		SFAM01.1	PPB	2/12/2025 11:48	SMPL
CCB062	CCB062	307	0.0162 -		SFAM01.1	PPB	2/12/2025 11:52	SMPL