

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134689

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.00	02/12/2025	16:20
2	CAL2	1	Water	NA	NA	20.2	7.01	02/12/2025	16:21
3	CAL3	1	Water	NA	NA	20.2	10.02	02/12/2025	16:25
4	ICV	1	Water	NA	NA	20.4	7.03	02/12/2025	16:33
5	CCV1	1	Water	NA	NA	20.2	2.01	02/12/2025	16:37
6	Q1357-01	1	Water	NA	NA	20.2	6.01	02/12/2025	16:40
7	Q1357-01DUP	1	Water	NA	NA	20.3	6.02	02/12/2025	16:41
8	Q1357-02	1	Water	NA	NA	20.6	6.10	02/12/2025	16:45
9	Q1357-04	1	Water	NA	NA	20.7	5.45	02/12/2025	16:55
10	Q1358-01	1	Water	NA	NA	20.9	6.89	02/12/2025	17:00
11	Q1358-02	1	Water	NA	NA	20.7	6.97	02/12/2025	17:10
12	Q1358-03	1	Water	NA	NA	20.4	7.07	02/12/2025	17:15
13	CCV2	1	Water	NA	NA	20.2	12.02	02/12/2025	17:20

WORKLIST(Hardcopy Internal Chain)

VB 134689

WorkList Name : ph w q1358 WorkList ID : 187683 Department : Wet-Chemistry Date : 02-12-2025 16:04:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1357-01	NWB-2135	Water	pH	Cool 4 deg C	PSEG03	N51	02/12/2025	9040C
Q1357-02	NWB-2136	Water	pH	Cool 4 deg C	PSEG03	N51	02/12/2025	9040C
Q1357-04	HCC-1	Water	pH	Cool 4 deg C	PSEG03	N51	02/12/2025	9040C
Q1358-01	207-A	Water	pH	Cool 4 deg C	PSEG03	N51	02/12/2025	9040C
Q1358-02	207-B	Water	pH	Cool 4 deg C	PSEG03	N51	02/12/2025	9040C
Q1358-03	207-C	Water	pH	Cool 4 deg C	PSEG03	N51	02/12/2025	9040C

Date/Time 02/12/25 16:15
 Raw Sample Received by: SP WOC / RM SM
 Raw Sample Relinquished by: RM SM

Date/Time 02/12/25 18:00
 Raw Sample Received by: RM SM
 Raw Sample Relinquished by: SP WOC