

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB134690

Slope : 98.4

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/12/2025	13:00
2	CAL2	1	Water	NA	NA	20.3	7.00	02/12/2025	13:02
3	CAL3	1	Water	NA	NA	20.3	10.03	02/12/2025	13:05
4	ICV	1	Water	NA	NA	20.2	7.02	02/12/2025	13:14
5	CCV1	1	Water	NA	NA	20.1	2.01	02/12/2025	13:15
6	Q1353-02	1	Solid	20.02	20	20.4	10.12	02/12/2025	13:25
7	Q1356-01	1	Solid	20.01	20	20.3	7.24	02/12/2025	13:30
8	Q1356-03	1	Solid	20.02	20	20.5	8.53	02/12/2025	13:37
9	Q1356-03DUP	1	Solid	20.03	20	20.6	8.54	02/12/2025	13:38
10	CCV2	1	Water	NA	NA	20.2	12.02	02/12/2025	13:44

WORKLIST(Hardcopy Internal Chain)

VB 134689

WorkList Name : ph w q1358 WorkList ID : 187683 Department : Wet-Chemistry Date : 02-12-2025 16:04:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1357-01	NWB-2135	Water	pH	Cool 4 deg C	PSEG03	N51	02/12/2025	9040C
Q1357-02	NWB-2136	Water	pH	Cool 4 deg C	PSEG03	N51	02/12/2025	9040C
Q1357-04	HCC-1	Water	pH	Cool 4 deg C	PSEG03	N51	02/12/2025	9040C
Q1358-01	207-A	Water	pH	Cool 4 deg C	PSEG03	N51	02/12/2025	9040C
Q1358-02	207-B	Water	pH	Cool 4 deg C	PSEG03	N51	02/12/2025	9040C
Q1358-03	207-C	Water	pH	Cool 4 deg C	PSEG03	N51	02/12/2025	9040C

Date/Time 02/12/25 16:15
 Raw Sample Received by: SP WOC / RM SM
 Raw Sample Relinquished by: RM SM

Date/Time 02/12/25 18:00
 Raw Sample Received by: RM SM
 Raw Sample Relinquished by: SP WOC