

LB134702


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

43007



A= 0.0000e+000

B= 2.3622e-004 slope

C= -5.3163e-002 y-intercept

Rho= 0.9996726

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.006	-0.006	201	0.000	201					0
0.05	0.050										0
0.20	0.200	0.218	0.018	1147	0.0 %	1147					1
2.50	2.500	2.516	0.016	10876	0.0 %	10876					1
5.00	5.000	5.061	0.061	21648	0.0 %	21648					1
7.50	7.500	7.305	-0.195	31150	0.0 %	31150					3
10.0	10.000	10.106	0.106	43007	0.0 %	43007					1

LB134702

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	201	-		0 SFAM01.1	PPB	2/13/2025 11:09	Std
	0.2 S01	1147	-		0.2 SFAM01.1	PPB	2/13/2025 11:12	Std
	2.5 S02	10876	-		2.5 SFAM01.1	PPB	2/13/2025 11:14	Std
	5 S03	21648	-		5 SFAM01.1	PPB	2/13/2025 11:16	Std
	7.5 S04	31150	-		7.5 SFAM01.1	PPB	2/13/2025 11:18	Std
	10 S05	43007	-		10 SFAM01.1	PPB	2/13/2025 11:32	Std
ICV003	ICV003	16926	3.9452 -		SFAM01.1	PPB	2/13/2025 11:37	SMPL
ICB003	ICB003	250	0.0059 -		SFAM01.1	PPB	2/13/2025 11:47	SMPL
CCV067	CCV067	21301	4.9786 -		SFAM01.1	PPB	2/13/2025 11:50	SMPL
CCB067	CCB067	211	-0.0033 -		SFAM01.1	PPB	2/13/2025 11:54	SMPL
PBW04	PBW04	238	0.0031 -		SFAM01.1	PPB	2/13/2025 11:59	SMPL
PBW05	PBW05	269	0.0104 -		SFAM01.1	PPB	2/13/2025 12:02	SMPL
MDL04	MDL04	432	0.0489 -		SFAM01.1	PPB	2/13/2025 12:04	SMPL
MDL05	MDL05	435	0.0496 -		SFAM01.1	PPB	2/13/2025 12:06	SMPL
CCV068	CCV068	21722	5.0781 -		SFAM01.1	PPB	2/13/2025 12:08	SMPL
CCB068	CCB068	212	-0.0031 -		SFAM01.1	PPB	2/13/2025 12:21	SMPL