

LB134706


7471B

INSTRUMENT ID: 1

Linear

 μ Abs.:

50962



A= 0.0000e+000

B= 1.9599e-004

C= -1.5541e-001

Rho= 0.9996037

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	-0.075	-0.075	412	0.000	412					
0.2	0.200	0.180	-0.020	1709	0.0 %	1709					10
2.5	2.500	2.536	0.036	13734	0.0 %	13734					1
5.0	5.000	5.070	0.070	26660	0.0 %	26660					1
7.5	7.500	7.657	0.157	39860	0.0 %	39860					2
10.0	10.000	9.833	-0.167	50962	0.0 %	50962					2

LB134706 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Feb 2025 13:11:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	412	-	0.0000	7471B	PPB	13 Feb 2025 13:26:39	S	Std
0.2 - 1	50.2	1709	-	0.2000	7471B	PPB	13 Feb 2025 13:28:56	S	Std
2.5 - 1	52.5	13734	-	2.5000	7471B	PPB	13 Feb 2025 13:31:13	S	Std
5.0 - 1	55	26660	-	5.0000	7471B	PPB	13 Feb 2025 13:33:29	S	Std
7.5 - 1	57.5	39860	-	7.5000	7471B	PPB	13 Feb 2025 13:35:48	S	Std
10.0 - 1	510	50962	-	10.0000	7471B	PPB	13 Feb 2025 13:40:52	S	Std
ICV04 - 1	ICV04	21161	3.9919	-7471B	PPB	13 Feb 2025 13:46:30	U	SMPL	
ICB04 - 1	ICB04	116	-0.1327	-7471B	PPB	13 Feb 2025 13:48:45	U	SMPL	
CCV69 - 1	CCV69	26438	5.0261	-7471B	PPB	13 Feb 2025 13:51:04	U	SMPL	
CCB69 - 1	CCB69	59	-0.1439	-7471B	PPB	13 Feb 2025 13:53:19	U	SMPL	
CRA - 1	CRA	1580	0.1542	-7471B	PPB	13 Feb 2025 14:00:50	U	SMPL	
HighStd - 1	HighStd	50925	9.8253	-7471B	PPB	13 Feb 2025 14:03:05	U	SMPL	
ChkStd - 1	ChkStd	32800	6.2730	-7471B	PPB	13 Feb 2025 14:05:20	U	SMPL	
PB166717BL - 1	PBS	568	-0.0441	-7471B	PPB	13 Feb 2025 14:10:51	U	SMPL	
PB166717BS - 1	LCSS	22710	4.2955	-7471B	PPB	13 Feb 2025 14:13:06	U	SMPL	
Q1346-01 - 1	SOIL-COMP	56056	10.8309	-7471B	PPB	13 Feb 2025 14:15:22	U	SMPL	
Q1346-01DUP - 1	SOIL-COMPDUP	52444	10.1230	-7471B	PPB	13 Feb 2025 14:17:43	U	SMPL	
Q1346-01MS - 1	SOIL-COMPMS	72996	14.1509	-7471B	PPB	13 Feb 2025 14:20:25	U	SMPL	
Q1346-01MSD - 1	SOIL-COMPMSD	72532	14.0600	-7471B	PPB	13 Feb 2025 14:25:53	U	SMPL	
Q1346-03 - 1	CONCRETE	7634	1.3408	-7471B	PPB	13 Feb 2025 14:28:11	U	SMPL	
CCV70 - 1	CCV70	26350	5.0089	-7471B	PPB	13 Feb 2025 14:30:48	U	SMPL	
CCB70 - 1	CCB70	101	-0.1356	-7471B	PPB	13 Feb 2025 14:33:06	U	SMPL	
Q1353-01 - 1	346	3825	0.5942	-7471B	PPB	13 Feb 2025 14:35:25	U	SMPL	
Q1354-01 - 1	NB-08-021125	2208	0.2773	-7471B	PPB	13 Feb 2025 14:37:41	U	SMPL	
Q1346-01LX5 - 1		9843	1.7737	-7471B	PPB	13 Feb 2025 14:39:58	U	SMPL	
Q1346-01A - 1		74294	14.4053	-7471B	PPB	13 Feb 2025 14:42:15	U	SMPL	
Q1346-01DLX2 - 1		27876	5.3080	-7471B	PPB	13 Feb 2025 14:44:33	U	SMPL	
Q1346-01DUPDLX2 - 1		25536	4.8493	-7471B	PPB	13 Feb 2025 14:47:10	U	SMPL	
Q1346-01MSDLX2 - 1		37477	7.1896	-7471B	PPB	13 Feb 2025 14:49:46	U	SMPL	
Q1346-01MSDDLX2 - 1		35689	6.8392	-7471B	PPB	13 Feb 2025 14:52:21	U	SMPL	
Q1346-01LDLX10 - 1		5246	0.8727	-7471B	PPB	13 Feb 2025 14:54:54	U	SMPL	
Q1346-01ADLX2 - 1		36865	7.0697	-7471B	PPB	13 Feb 2025 14:57:26	U	SMPL	
CCV71 - 1	CCV71	26671	5.0718	-7471B	PPB	13 Feb 2025 14:59:45	U	SMPL	
CCB71 - 1	CCB71	-218	-0.1981	-7471B	PPB	13 Feb 2025 15:02:08	U	SMPL	