

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB134707

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3158

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	02/13/2025	09:40
2	Q1356-04		1	14	100.00	>212.0	>212.0	02/13/2025	10:10
3	Q1356-04DUP		1	14	100.00	>212.0	>212.0	02/13/2025	10:40
4	Q1356-05		1	14	100.00	>212.0	>212.0	02/13/2025	11:10
5	Q1356-06		1	12	100.00	>212.0	>212.0	02/13/2025	11:40
6	Q1356-07		1	11	100.00	>212.0	>212.0	02/13/2025	12:10
7	Q1356-08		1	14	100.00	>212.0	>212.0	02/13/2025	12:40
8	Q1356-09		1	11	100.00	>212.0	>212.0	02/13/2025	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6134707

WorkList Name : fp-02-12

WorkList ID : 187682

Department : Wet-Chemistry

Date : 02-12-2025 15:01:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1356-04 J	CARBON-WATER	Water	Flash Point	Cool 4 deg C	PSEG04	N51	02/11/2025	1010B
Q1356-05 J	CARBON-FB	Water	Flash Point	Cool 4 deg C	PSEG04	N51	02/11/2025	1010B
Q1356-06 H	WATER-A	Water	Flash Point	Cool 4 deg C	PSEG04	N51	02/11/2025	1010B
Q1356-07 L	WATER-B	Water	Flash Point	Cool 4 deg C	PSEG04	N51	02/11/2025	1010B
Q1356-08 J	WATER-FB	Water	Flash Point	Cool 4 deg C	PSEG04	N51	02/11/2025	1010B
Q1356-09 J	SOIL-FB	Water	Flash Point	Cool 4 deg C	PSEG04	N51	02/11/2025	1010B

Date/Time 02/13/2025 09:25
 Raw Sample Received by: RM LWC
 Raw Sample Relinquished by: JH LWC

Date/Time 02/13/2025 13:50
 Raw Sample Received by: JH LWC
 Raw Sample Relinquished by: RM LWC