

LB134718

SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41634

A= 0.0000e+000

B= 2.4171e-004 slope

C= -8.4461e-002 intercept

Rho= 0.9999244

Accept=Accepted

Max. Conc: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% (1)
0.00	0.000	-0.019	-0.019	271	0.000	271					
0.05	0.050										
0.20	0.200	0.172	-0.028	1061	0.0 %	1061					-14
2.50	2.500	2.514	0.014	10750	0.0 %	10750					1
5.00	5.000	5.095	0.095	21427	0.0 %	21427					2
7.50	7.500	7.460	-0.040	31212	0.0 %	31212					-1
10.0	10.000	9.979	-0.021	41634	0.0 %	41634					0

LB134718

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	271	-		0 SFAM01.1	PPB	2/14/2025 8:36	Std
	0.2 S01	1061	-		0.2 SFAM01.1	PPB	2/14/2025 8:38	Std
	2.5 S02	10750	-		2.5 SFAM01.1	PPB	2/14/2025 8:41	Std
	5 S03	21427	-		5 SFAM01.1	PPB	2/14/2025 8:43	Std
	7.5 S04	31212	-		7.5 SFAM01.1	PPB	2/14/2025 8:45	Std
	10 S05	43518	-		10 SFAM01.1	PPB	2/14/2025 8:48	Std
ICV005	ICV005	17605	4.1708 -		SFAM01.1	PPB	2/14/2025 8:53	SMPL
ICB005	ICB005	271	-0.019 -		SFAM01.1	PPB	2/14/2025 8:58	SMPL
CCV072	CCV072	21346	5.075 -		SFAM01.1	PPB	2/14/2025 9:01	SMPL
CCB072	CCB072	375	0.0062 -		SFAM01.1	PPB	2/14/2025 9:05	SMPL
PBW06	PBW06	292	-0.0139 -		SFAM01.1	PPB	2/14/2025 9:08	SMPL
PBW07	PBW07	319	-0.0074 -		SFAM01.1	PPB	2/14/2025 9:10	SMPL
MDL06	MDL06	567	0.0526 -		SFAM01.1	PPB	2/14/2025 9:17	SMPL
MDL07	MDL07	554	0.0494 -		SFAM01.1	PPB	2/14/2025 9:20	SMPL
CCV073	CCV073	21718	5.1649 -		SFAM01.1	PPB	2/14/2025 9:22	SMPL
CCB073	CCB073	270	-0.0192 -		SFAM01.1	PPB	2/14/2025 9:27	SMPL