

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134720

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.4	02/14/2025	13:40
2	Q1357-03		1	11	100.00	>212.0	>212.0	02/14/2025	14:10
3	Q1357-03DUP		1	11	100.00	>212.0	>212.0	02/14/2025	14:40
4	Q1357-04		1	10	63.00	145.4	144.4	02/14/2025	15:10
5	Q1358-04		1	14	100.00	>212.0	>212.0	02/14/2025	15:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

Lb 134720

WorkList Name : FP-2-14

WorkList ID : 187713

Department : Wet-Chemistry

Date : 02-14-2025 08:46:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1357-03	B NWB-2135-36	Water	Flash Point	Cool 4 deg C	PSEG03	N51	02/12/2025	1010B
Q1357-04	B HCC-1	Water	Flash Point	Cool 4 deg C	PSEG03	N51	02/12/2025	1010B
Q1358-04	B 207-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	N51	02/12/2025	1010B

Date/Time 02/14/2025 13.30
 Raw Sample Received by: RM (wlc)
 Raw Sample Relinquished by: RM (wlc)

Date/Time 02/14/2025 16.20
 Raw Sample Received by: RM (wlc)
 Raw Sample Relinquished by: RM (wlc)