

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134726

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	02/17/2025	10:45
2	Q1371-01		1	15	100.00	>212.0	>212.0	02/17/2025	11:15
3	Q1371-01DUP		1	15	100.00	>212.0	>212.0	02/17/2025	11:45
4	Q1371-02		1	14	100.00	>212.0	>212.0	02/17/2025	12:15
5	Q1371-03		1	15	100.00	>212.0	>212.0	02/17/2025	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6134726

WorkList Name : fp-02-17 WorkList ID : 187723 Department : Wet-Chemistry Date : 02-17-2025 07:54:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1371-01 B	AUD-25-0011	Water	Flash Point	Cool 4 deg C	PSEG03	H21	02/14/2025	1010B
Q1371-02 B	AUD-25-0012	Water	Flash Point	Cool 4 deg C	PSEG03	H21	02/14/2025	1010B
Q1371-03 B	AUD-25-0013	Water	Flash Point	Cool 4 deg C	PSEG03	H21	02/14/2025	1010B

Date/Time 02/17/2025 10:30
 Raw Sample Received by: RM [signature]
 Raw Sample Relinquished by: [signature]

Date/Time 02/17/2025 13:30
 Raw Sample Received by: [signature]
 Raw Sample Relinquished by: RM [signature]