

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : rubina

Run Number: LB134729

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/17/2025	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	02/17/2025	08:11
3	CAL3	1	Water	NA	NA	20.1	10.03	02/17/2025	08:12
4	ICV	1	Water	NA	NA	20.3	7.01	02/17/2025	08:15
5	CCV1	1	Water	NA	NA	20.2	2.01	02/17/2025	08:22
6	Q1371-01	1	Water	NA	NA	20.4	4.93	02/17/2025	08:30
7	Q1371-01DUP	1	Water	NA	NA	20.5	4.94	02/17/2025	08:31
8	Q1371-02	1	Water	NA	NA	20.9	6.62	02/17/2025	08:37
9	Q1371-03	1	Water	NA	NA	20.2	5.96	02/17/2025	08:44
10	CCV2	1	Water	NA	NA	20.2	12.02	02/17/2025	08:45

WORKLIST(Hardcopy Internal Chain)

134229

WorkList Name : ph q1371

WorkList ID : 187726

Department : Wet-Chemistry

Date : 02-17-2025 07:56:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1371-01	AUD-25-0011	Water	pH	Cool 4 deg C	PSEG03	H21	02/14/2025	9040C
Q1371-02	AUD-25-0012	Water	pH	Cool 4 deg C	PSEG03	H21	02/14/2025	9040C
Q1371-03	AUD-25-0013	Water	pH	Cool 4 deg C	PSEG03	H21	02/14/2025	9040C

Date/Time 02/17/25 08:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 02/17/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]