

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : rubina

Run Number: LB134730

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/17/2025	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	02/17/2025	08:11
3	CAL3	1	Water	NA	NA	20.1	10.03	02/17/2025	08:12
4	ICV	1	Water	NA	NA	20.3	7.01	02/17/2025	08:15
5	CCV1	1	Water	NA	NA	20.1	2.02	02/17/2025	08:55
6	Q1372-01	1	Solid	20.02	20	22.2	7.85	02/17/2025	09:10
7	Q1372-01DUP	1	Solid	20.03	20	22.1	7.86	02/17/2025	09:11
8	CCV2	1	Water	NA	NA	20.3	12.02	02/17/2025	09:15

WORKLIST(Hardcopy Internal Chain)

UB 134730

WorkList Name : ph q1372 s

WorkList ID : 187727

Department : Wet-Chemistry

Date : 02-17-2025 07:56:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1372-01	AU-05-021425	Solid	pH	Cool 4 deg C	PSEG05	H21	02/14/2025	9045D

Date/Time 02/17/25 08:00

Raw Sample Received by: JH (CSC)

Raw Sample Relinquished by: JH (CSC)

Date/Time 02/17/25 02:17:25 12:00

Raw Sample Received by: JH (CSC)

Raw Sample Relinquished by: JH (CSC)