

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : rubina

Run Number: LB134733

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/17/2025	13:35
2	CAL2	1	Water	NA	NA	20.2	7.00	02/17/2025	13:36
3	CAL3	1	Water	NA	NA	20.3	10.02	02/17/2025	13:39
4	ICV	1	Water	NA	NA	20.1	7.00	02/17/2025	13:44
5	CCV1	1	Water	NA	NA	20.2	2.01	02/17/2025	13:50
6	Q1373-01	1	Solid	20.02	20	20.5	6.77	02/17/2025	14:00
7	Q1373-01DUP	1	Solid	20.03	20	20.6	6.78	02/17/2025	14:01
8	Q1373-02	1	Solid	20.04	20	21.6	8.16	02/17/2025	14:15
9	CCV2	1	Water	NA	NA	20.3	12.02	02/17/2025	14:20

WORKLIST(Hardcopy Internal Chain)

WJ B4433

WorkList Name : corrosivity q1337 WorkList ID : 187738 Department : Wet-Chemistry Date : 02-17-2025 13:13:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1373-01	VB27161	Solid	Corrosivity	Cool 4 deg C	PSEG03	H21	02/14/2025	9045D
Q1373-02	VDR25992	Solid	Corrosivity	Cool 4 deg C	PSEG03	H21	02/14/2025	9045D

Date/Time 02/17/25 13:23
Raw Sample Received by: AB Gage
Raw Sample Relinquished by: RS (E4-1000)

Date/Time 02/14/25 16:00
Raw Sample Received by: RS (E4-1000)
Raw Sample Relinquished by: AB Gage