

LB134741


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43234



A= 0.0000e+000

B= 2.2697e-004 *slope*C= -1.1753e-002 *+intercept*

Rho= 0.9993782

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	G/D
0.00	0.000	-0.023	-0.023	-49	0.000	-49					=
0.05	0.050										
0.20	0.200	0.181	-0.019	848	0.0 %	848					-16
2.50	2.500	2.469	-0.031	10930	0.0 %	10930					-1
5.00	5.000	5.029	0.029	22207	0.0 %	22207					1
7.50	7.500	7.744	0.244	34169	0.0 %	34169					3
10.0	10.000	9.801	-0.199	43234	0.0 %	43234					-2

LB134741
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-49	-		0 SFAM01.1	PPB	2/18/2025 13:48	Std
	0.2 S01	848	-		0.2 SFAM01.1	PPB	2/18/2025 13:50	Std
	2.5 S02	10930	-		2.5 SFAM01.1	PPB	2/18/2025 13:53	Std
	5 S03	22207	-		5 SFAM01.1	PPB	2/18/2025 13:55	Std
	7.5 S04	34169	-		7.5 SFAM01.1	PPB	2/18/2025 13:57	Std
	10 S05	43234	-		10 SFAM01.1	PPB	2/18/2025 14:02	Std
ICV008	ICV008	16141	3.6518 -		SFAM01.1	PPB	2/18/2025 14:06	SMPL
ICB008	ICB008	-342	-0.0894 -		SFAM01.1	PPB	2/18/2025 14:08	SMPL
CCV081	CCV081	22369	5.0653 -		SFAM01.1	PPB	2/18/2025 14:10	SMPL
CCB081	CCB081	-329	-0.0864 -		SFAM01.1	PPB	2/18/2025 14:13	SMPL
PB166766BL	PBW766	-40	-0.0208 -		SFAM01.1	PPB	2/18/2025 14:15	SMPL
Q1200-01	ME2964	-20	-0.0163 -		SFAM01.1	PPB	2/18/2025 14:17	SMPL
Q1200-02	ME2964D	-65	-0.0265 -		SFAM01.1	PPB	2/18/2025 14:19	SMPL
Q1200-03	ME2964S	4648	1.0432 -		SFAM01.1	PPB	2/18/2025 14:22	SMPL
Q1200-04	ME2981	-155	-0.0469 -		SFAM01.1	PPB	2/18/2025 14:24	SMPL
Q1200-05	ME2982	-92	-0.0326 -		SFAM01.1	PPB	2/18/2025 14:26	SMPL
Q1200-06	ME2983	-66	-0.0267 -		SFAM01.1	PPB	2/18/2025 14:28	SMPL
Q1200-07	ME2984	-92	-0.0326 -		SFAM01.1	PPB	2/18/2025 14:31	SMPL
Q1200-08	ME2985	-94	-0.0331 -		SFAM01.1	PPB	2/18/2025 14:33	SMPL
Q1200-09	ME2986	-35	-0.0197 -		SFAM01.1	PPB	2/18/2025 14:35	SMPL
Q1200-10	ME2987	-246	-0.0676 -		SFAM01.1	PPB	2/18/2025 14:38	SMPL
Q1200-11	ME2988	5	-0.0106 -		SFAM01.1	PPB	2/18/2025 14:40	SMPL
Q1200-12	ME2992	29	-0.0052 -		SFAM01.1	PPB	2/18/2025 14:42	SMPL
Q1200-13	ME2994	-55	-0.0242 -		SFAM01.1	PPB	2/18/2025 14:44	SMPL
Q1200-14	ME2995	-49	-0.0229 -		SFAM01.1	PPB	2/18/2025 14:47	SMPL
Q1200-15	ME2999	-16	-0.0154 -		SFAM01.1	PPB	2/18/2025 14:49	SMPL
Q1200-16	ME2997	-54	-0.024 -		SFAM01.1	PPB	2/18/2025 14:51	SMPL
Q1200-17	ME2998	-25	-0.0174 -		SFAM01.1	PPB	2/18/2025 14:53	SMPL
Q1200-18	ME29A0	-43	-0.0215 -		SFAM01.1	PPB	2/18/2025 14:56	SMPL
Q1200-19	ME29A1	2	-0.0113 -		SFAM01.1	PPB	2/18/2025 14:58	SMPL
CCV082	CCV082	22378	5.0674 -		SFAM01.1	PPB	2/18/2025 15:00	SMPL
CCB082	CCB082	-255	-0.0696 -		SFAM01.1	PPB	2/18/2025 15:03	SMPL
Q1200-20	ME29A2	0	-0.0118 -		SFAM01.1	PPB	2/18/2025 15:05	SMPL
Q1200-21	ME29A3	-10	-0.014 -		SFAM01.1	PPB	2/18/2025 15:07	SMPL
Q1223-17	A6310	-70	-0.0276 -		SFAM01.1	PPB	2/18/2025 15:09	SMPL
PB166767BL	PBW767	330	0.0631 -		SFAM01.1	PPB	2/18/2025 15:12	SMPL
Q1204-01	ME2975	305	0.0575 -		SFAM01.1	PPB	2/18/2025 15:14	SMPL
Q1204-02	ME2975D	264	0.0482 -		SFAM01.1	PPB	2/18/2025 15:16	SMPL
Q1204-03	ME2975S	5854	1.3169 -		SFAM01.1	PPB	2/18/2025 15:18	SMPL
Q1204-04	ME2978	324	0.0618 -		SFAM01.1	PPB	2/18/2025 15:21	SMPL
Q1204-05	ME2979	470	0.0949 -		SFAM01.1	PPB	2/18/2025 15:23	SMPL
Q1204-06	ME2989	519	0.106 -		SFAM01.1	PPB	2/18/2025 15:25	SMPL
Q1204-07	ME2991	481	0.0974 -		SFAM01.1	PPB	2/18/2025 15:28	SMPL
Q1204-08	ME2996	4700	1.055 -		SFAM01.1	PPB	2/18/2025 15:30	SMPL
Q1204-09	ME29A4	70	0.0041 -		SFAM01.1	PPB	2/18/2025 15:32	SMPL
Q1204-10	ME29A5	484	0.0981 -		SFAM01.1	PPB	2/18/2025 15:34	SMPL

LB134741

INSTRUMENT ID : CV1

Q1204-11	ME29A7	329	0.0629 -	SFAM01.1	PPB	2/18/2025 15:37 SMPL
Q1204-12	ME29A8	290	0.0541 -	SFAM01.1	PPB	2/18/2025 15:39 SMPL
Q1204-13	ME29B0	379	0.0743 -	SFAM01.1	PPB	2/18/2025 15:41 SMPL
Q1204-14	ME29B1	413	0.082 -	SFAM01.1	PPB	2/18/2025 15:44 SMPL
Q1204-15	ME29A9	337	0.0647 -	SFAM01.1	PPB	2/18/2025 15:46 SMPL
Q1204-16	ME29B2	340	0.0654 -	SFAM01.1	PPB	2/18/2025 15:48 SMPL
CCV083	CCV083	21566	4.8831 -	SFAM01.1	PPB	2/18/2025 15:50 SMPL
CCB083	CCB083	-238	-0.0658 -	SFAM01.1	PPB	2/18/2025 15:53 SMPL
Q1204-17	ME29B3	320	0.0609 -	SFAM01.1	PPB	2/18/2025 15:55 SMPL
Q1204-18	ME29B4	293	0.0547 -	SFAM01.1	PPB	2/18/2025 15:57 SMPL
Q1204-19	ME29B5	355	0.0688 -	SFAM01.1	PPB	2/18/2025 15:59 SMPL
Q1204-20	ME29B8	317	0.0602 -	SFAM01.1	PPB	2/18/2025 16:02 SMPL
Q1204-21	ME29B9	360	0.07 -	SFAM01.1	PPB	2/18/2025 16:04 SMPL
Q1204-22	ME29C0	366	0.0713 -	SFAM01.1	PPB	2/18/2025 16:06 SMPL
CCV084	CCV084	21765	4.9282 -	SFAM01.1	PPB	2/18/2025 16:09 SMPL
CCB084	CCB084	-283	-0.076 -	SFAM01.1	PPB	2/18/2025 16:11 SMPL