

LB134745


7470A

INSTRUMENT ID: C82

Linear

μ Abs.:

36612



A= 0.0000e+000

B= 2.5888e-004

C= 4.1130e-002

Rho= 0.9958632

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.055	0.055	54	0.000	54	0				15
0.2	0.200	0.229	0.029	727	0.0 %	727					-16
2.5	2.500	2.259	-0.241	8566	0.0 %	8566					0
5.0	5.000	5.021	0.021	19235	0.0 %	19235					5
7.5	7.500	8.117	0.617	31196	0.0 %	31196					
10.0	10.000	9.519	-0.481	36612	0.0 %	36612					

LB134745 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Feb 2025 11:12:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	54	-	0.0000	7470A	PPB	19 Feb 2025 11:18:12	S	Std
0.2 - 1	50-2	727	-	0.2000	7470A	PPB	19 Feb 2025 11:20:29	S	Std
2.5 - 1	52.5	8566	-	2.5000	7470A	PPB	19 Feb 2025 11:22:47	S	Std
5.0 - 1	55	19235	-	5.0000	7470A	PPB	19 Feb 2025 11:25:04	S	Std
7.5 - 1	57.5	31196	-	7.5000	7470A	PPB	19 Feb 2025 11:30:02	S	Std
10.0 - 1	5W	36612	-	10.0000	7470A	PPB	19 Feb 2025 11:35:00	S	Std
ICV09 - 1	ICV09	15039	3.9344	-7470A	PPB	PPB	19 Feb 2025 11:40:30	U	SMPL
ICB09 - 1	ICB09	-179	-0.0052	-7470A	PPB	PPB	19 Feb 2025 11:42:45	U	SMPL
CCV85 - 1	CCV85	19738	5.1509	-7470A	PPB	PPB	19 Feb 2025 11:45:02	U	SMPL
CCB85 - 1	CCB85	-269	-0.0285	-7470A	PPB	PPB	19 Feb 2025 11:47:18	U	SMPL
CRA - 1	CRA	737	0.2319	-7470A	PPB	PPB	19 Feb 2025 11:49:35	U	SMPL
HighStd - 1	HighStd	36733	9.5505	-7470A	PPB	PPB	19 Feb 2025 11:51:50	U	SMPL
ChkStd - 1	ChkStd	27620	7.1913	-7470A	PPB	PPB	19 Feb 2025 11:54:06	U	SMPL
PB166769BL - 1	PBW	-398	-0.0619	-7470A	PPB	PPB	19 Feb 2025 11:56:24	U	SMPL
Q1373-01 - 1	VB27161	-243	-0.0218	-7470A	PPB	PPB	19 Feb 2025 12:00:59	U	SMPL
Q1373-02 - 1	VDR25992	118	0.0717	-7470A	PPB	PPB	19 Feb 2025 12:08:26	U	SMPL
Q1373-02DUP - 1	VDR25992DUP	91	0.0647	-7470A	PPB	PPB	19 Feb 2025 12:10:43	U	SMPL
Q1373-02MS - 1	VDR25992MS	18599	4.8560	-7470A	PPB	PPB	19 Feb 2025 12:13:00	U	SMPL
Q1373-02MSD - 1	VDR25992MSD	17164	4.4845	-7470A	PPB	PPB	19 Feb 2025 12:15:15	U	SMPL
CCV86 - 1	CCV86	20915	5.4556	-7470A	PPB	PPB	19 Feb 2025 12:17:33	U	SMPL
CCB86 - 1	CCB86	-239	-0.0207	-7470A	PPB	PPB	19 Feb 2025 12:19:50	U	SMPL
PB166735TB - 1		25	0.0476	-7470A	PPB	PPB	19 Feb 2025 12:22:08	U	SMPL
Q1373-02LX5 - 1		-29	0.0336	-7470A	PPB	PPB	19 Feb 2025 12:24:23	U	SMPL
Q1373-02A - 1		16159	4.2243	-7470A	PPB	PPB	19 Feb 2025 12:26:38	U	SMPL
PB166769BS - 1	LCSW	15348	4.0144	-7470A	PPB	PPB	19 Feb 2025 12:28:54	U	SMPL
CCV87 - 1	CCV87	21084	5.4993	-7470A	PPB	PPB	19 Feb 2025 12:31:12	U	SMPL
CCB87 - 1	CCB87	-243	-0.0218	-7470A	PPB	PPB	19 Feb 2025 12:33:30	U	SMPL