

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB134749

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/19/2025	15:10
2	CAL2	1	Water	NA	NA	20.2	7.00	02/19/2025	15:11
3	CAL3	1	Water	NA	NA	20.1	10.03	02/19/2025	15:15
4	ICV	1	Water	NA	NA	20.2	7.01	02/19/2025	15:20
5	CCV1	1	Water	NA	NA	20.3	2.01	02/19/2025	15:25
6	Q1393-05	1	Water	NA	NA	20.7	6.90	02/19/2025	15:35
7	Q1393-05DUP	1	Water	NA	NA	20.8	6.91	02/19/2025	15:36
8	Q1393-06	1	Water	NA	NA	20.9	7.00	02/19/2025	15:44
9	Q1393-07	1	Water	NA	NA	20.1	6.86	02/19/2025	15:50
10	CCV2	1	Water	NA	NA	20.2	12.02	02/19/2025	16:00

WORKLIST(Hardcopy Internal Chain)

UP 134749

WorkList Name : ph q1393

WorkList ID : 187770

Department : Wet-Chemistry

Date : 02-19-2025 14:42:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1393-05	PURGE-WATER-COMP	Water	pH	Cool 4 deg C	PSEG03	H21	02/19/2025	9040C
Q1393-06	PURGE-WATER-1	Water	pH	Cool 4 deg C	PSEG03	H21	02/19/2025	9040C
Q1393-07	PURGE-WATER-2	Water	pH	Cool 4 deg C	PSEG03	H21	02/19/2025	9040C

Date/Time 02/19/25 14:55

Raw Sample Received by: 48 WDC

Raw Sample Relinquished by: ST (SM)

Date/Time 02/19/25

Raw Sample Received by: ST (SM)

Raw Sample Relinquished by: 48 WDC