

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134754

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	02/20/2025	08:35
2	Q1390-01		1	15	100.00	>212.0	>212.0	02/20/2025	09:05
3	Q1392-01		1	14	100.00	>212.0	>212.0	02/20/2025	09:35
4	Q1393-05		1	15	100.00	>212.0	>212.0	02/20/2025	10:05
5	Q1393-05DUP		1	15	100.00	>212.0	>212.0	02/20/2025	10:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66134754

WorkList Name : FP-2-19

WorkList ID : 187779

Department : Wet-Chemistry

Date : 02-19-2025 15:57:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1390-01B	3771	Water	Flash Point	Cool 4 deg C	PSEG03	H21	02/19/2025	1010B
Q1392-01B	435	Water	Flash Point	Cool 4 deg C	PSEG03	H21	02/19/2025	1010B
Q1393-05 T	PURGE-WATER-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	H21	02/19/2025	1010B

Date/Time 02/20/2025 08:20
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: RM CWG

Date/Time 02/20/2025 11:15
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: RM CWG