

L13134771


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

56791



A= 0.0000e+000

B= 1.7846e-004

C= -2.8325e-002

Rho= 0.9996331

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/13
0.0	0.000	-0.028	-0.028	0	0.000	0					
0.2	0.200	0.224	0.024	1415	0.0 %	1415					
2.5	2.500	2.543	0.043	14409	0.0 %	14409					
5.0	5.000	5.059	0.059	28505	0.0 %	28505					
7.5	7.500	7.295	-0.205	41038	0.0 %	41038					
10.0	10.000	10.107	0.107	56791	0.0 %	56791					

LB134771 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Feb 2025 12:17:22

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	0	-	0.0000	7471B	PPB	21 Feb 2025 12:29:59	S	Std
0.2 - 1	50.2	1415	-	0.2000	7471B	PPB	21 Feb 2025 12:32:16	S	Std
2.5 - 1	52.5	14409	-	2.5000	7471B	PPB	21 Feb 2025 12:34:33	S	Std
5.0 - 1	55	28505	-	5.0000	7471B	PPB	21 Feb 2025 12:36:50	S	Std
7.5 - 1	57.5	41038	-	7.5000	7471B	PPB	21 Feb 2025 12:39:09	S	Std
10.0 - 1	60	56791	-	10.0000	7471B	PPB	21 Feb 2025 12:41:28	S	Std
ICV13 - 1	ICV13	22577	4.0008	-	7471B	PPB	21 Feb 2025 12:52:03	U	SMPL
ICB13 - 1	ICB13	-319	-0.0853	-	7471B	PPB	21 Feb 2025 12:54:19	U	SMPL
CCV01 - 1	CCV01	26683	4.7336	-	7471B	PPB	21 Feb 2025 12:56:37	U	SMPL
CCB01 - 1	CCB01	1	-0.0281	-	7471B	PPB	21 Feb 2025 13:01:50	U	SMPL
CRA - 1	CRA	1331	0.2092	-	7471B	PPB	21 Feb 2025 13:04:04	U	SMPL
HighStd - 1	HighStd	56305	10.0200	-	7471B	PPB	21 Feb 2025 13:06:19	U	SMPL
ChkStd - 1	ChkStd	37085	6.5900	-	7471B	PPB	21 Feb 2025 13:08:34	U	SMPL
PB166828BL - 1	PB166828BL	-2	-0.0287	-	7471B	PPB	21 Feb 2025 13:13:49	U	SMPL
PB166828BS - 1	PB166828BS	21230	3.7604	-	7471B	PPB	21 Feb 2025 13:16:05	U	SMPL
Q1383-01 - 1	OU4-PCS-TC-11-021725	1448	0.2301	-	7471B	PPB	21 Feb 2025 13:18:21	U	SMPL
Q1383-03 - 1	OU4-PCS-TC-12-021725	954	0.1419	-	7471B	PPB	21 Feb 2025 13:20:40	U	SMPL
Q1383-05 - 1	OU4-PCS-TC-13-021725	327	0.0300	-	7471B	PPB	21 Feb 2025 13:23:00	U	SMPL
Q1383-07 - 1	OU4-PCS-TC-14-021725	258	0.0177	-	7471B	PPB	21 Feb 2025 13:25:18	U	SMPL
Q1383-09 - 1	OU4-PCS-TC-15-021725	178	0.0034	-	7471B	PPB	21 Feb 2025 13:27:34	U	SMPL
CCV02 - 1	CCV02	26314	4.6678	-	7471B	PPB	21 Feb 2025 13:29:49	U	SMPL
CCB02 - 1	CCB02	-351	-0.0910	-	7471B	PPB	21 Feb 2025 13:32:05	U	SMPL
Q1383-11 - 1	OU4-PCS-TC-16-021725	791	0.1128	-	7471B	PPB	21 Feb 2025 13:34:24	U	SMPL
Q1383-13 - 1	OU4-PCS-TC-17-021725	230	0.0127	-	7471B	PPB	21 Feb 2025 13:36:39	U	SMPL
Q1383-15 - 1	OU4-PCS-TC-18-021725	188	0.0052	-	7471B	PPB	21 Feb 2025 13:38:55	U	SMPL
Q1383-17 - 1	OU4-PCS-TC-19-021725	223	0.0115	-	7471B	PPB	21 Feb 2025 13:41:11	U	SMPL
Q1383-19 - 1	OU4-PCS-TC-20-021725	192	0.0059	-	7471B	PPB	21 Feb 2025 13:43:27	U	SMPL
Q1383-21 - 1	OU4-CF-15-021725	216	0.0102	-	7471B	PPB	21 Feb 2025 13:45:43	U	SMPL
Q1397-01 - 1	STOCK-PILE	4561	0.7856	-	7471B	PPB	21 Feb 2025 13:47:59	U	SMPL
Q1397-03 - 1	TP7	4983	0.8610	-	7471B	PPB	21 Feb 2025 13:50:16	U	SMPL
Q1397-05 - 1	TP8	6093	1.0591	-	7471B	PPB	21 Feb 2025 13:52:34	U	SMPL
Q1397-05DUP - 1	TP8DUP	7472	1.3052	-	7471B	PPB	21 Feb 2025 13:54:51	U	SMPL
CCV03 - 1	CCV03	27762	4.9262	-	7471B	PPB	21 Feb 2025 13:57:08	U	SMPL
CCB03 - 1	CCB03	6	-0.0273	-	7471B	PPB	21 Feb 2025 14:02:17	U	SMPL
Q1397-05MSD - 1	TP8MSD	25766	4.5700	-	7471B	PPB	21 Feb 2025 14:07:16	U	SMPL
Q1397-05MS - 1	TP8MS	26569	4.7133	-	7471B	PPB	21 Feb 2025 14:10:08	U	SMPL
PB166830BL - 1	PB166830BL	-84	-0.0433	-	7471B	PPB	21 Feb 2025 14:12:51	U	SMPL
PB166830BS - 1	PB166830BS	21443	3.7985	-	7471B	PPB	21 Feb 2025 14:15:08	U	SMPL
Q1382-01 - 1	OU4-PCS-TC-01-021725	1203	0.1864	-	7471B	PPB	21 Feb 2025 14:17:25	U	SMPL
Q1382-03 - 1	OU4-PCS-TC-02-021725	323	0.0293	-	7471B	PPB	21 Feb 2025 14:19:44	U	SMPL
Q1382-05 - 1	OU4-PCS-TC-03-021725	87	-0.0128	-	7471B	PPB	21 Feb 2025 14:22:01	U	SMPL
Q1382-07 - 1	OU4-PCS-TC-04-021725	164	0.0009	-	7471B	PPB	21 Feb 2025 14:24:18	U	SMPL
Q1382-09 - 1	OU4-PCS-TC-05-021725	272	0.0202	-	7471B	PPB	21 Feb 2025 14:26:36	U	SMPL
Q1382-11 - 1	OU4-PCS-TC-06-021725	256	0.0174	-	7471B	PPB	21 Feb 2025 14:28:51	U	SMPL
CCV04 - 1	CCV04	27876	4.9465	-	7471B	PPB	21 Feb 2025 14:31:06	U	SMPL
CCB04 - 1	CCB04	-346	-0.0901	-	7471B	PPB	21 Feb 2025 14:33:22	U	SMPL
Q1382-13 - 1	OU4-PCS-TC-07-021725	310	0.0270	-	7471B	PPB	21 Feb 2025 14:35:41	U	SMPL
Q1382-15 - 1	OU4-PCS-TC-08-021725	183	0.0043	-	7471B	PPB	21 Feb 2025 14:37:57	U	SMPL
Q1382-17 - 1	OU4-PCS-TC-09-021725	159	0.0001	-	7471B	PPB	21 Feb 2025 14:40:13	U	SMPL
Q1382-19 - 1	OU4-PCS-TC-10-021725	188	0.0052	-	7471B	PPB	21 Feb 2025 14:42:29	U	SMPL
Q1384-01 - 1	OR-03-021823025	1986	0.3261	-	7471B	PPB	21 Feb 2025 14:44:45	U	SMPL
Q1385-01 - 1	ST-1	235	0.0136	-	7471B	PPB	21 Feb 2025 14:47:01	U	SMPL
Q1388-01 - 1	72-12016	1339	0.2106	-	7471B	PPB	21 Feb 2025 14:49:18	U	SMPL
Q1388-01DUP - 1	72-12016DUP	1277	0.1996	-	7471B	PPB	21 Feb 2025 14:54:24	U	SMPL
Q1388-01MS - 1	72-12016MS	20796	3.6830	-	7471B	PPB	21 Feb 2025 14:56:40	U	SMPL
Q1388-01MSD - 1	72-12016MSD	23287	4.1275	-	7471B	PPB	21 Feb 2025 14:58:56	U	SMPL
CCV05 - 1	CCV05	30505	5.4157	-	7471B	PPB	21 Feb 2025 15:01:14	U	SMPL
CCB05 - 1	CCB05	3	-0.0278	-	7471B	PPB	21 Feb 2025 15:06:22	U	SMPL
Q1393-01 - 1	SOIL-COMP	401	0.0432	-	7471B	PPB	21 Feb 2025 15:08:37	U	SMPL
Q1393-03 - 1	OD-281	620	0.0823	-	7471B	PPB	21 Feb 2025 15:10:53	U	SMPL
Q1397-05LX5 - 1		1500	0.2394	-	7471B	PPB	21 Feb 2025 15:13:08	U	SMPL
Q1397-05A - 1		26387	4.6808	-	7471B	PPB	21 Feb 2025 15:15:24	U	SMPL
Q1388-01LX5 - 1		-290	-0.0801	-	7471B	PPB	21 Feb 2025 15:17:40	U	SMPL
Q1388-01A - 1		23245	4.1200	-	7471B	PPB	21 Feb 2025 15:20:01	U	SMPL
CCV06 - 1	CCV06	27456	4.8716	-	7471B	PPB	21 Feb 2025 15:22:20	U	SMPL
CCB06 - 1	CCB06	-401	-0.0999	-	7471B	PPB	21 Feb 2025 15:24:39	U	SMPL
Q1388-01DUP - 1	72-12016DUP	2143	0.3541	-	7471B	PPB	21 Feb 2025 14:51:35	U	SMPL