

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134783

Slope : 98.6

BalanceID:

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/24/2025	08:10
2	CAL2	1	Water	NA	NA	20.2	7.00	02/24/2025	08:11
3	CAL3	1	Water	NA	NA	20.2	10.02	02/24/2025	08:12
4	ICV	1	Water	NA	NA	20.3	7.00	02/24/2025	08:15
5	CCV1	1	Water	NA	NA	20.3	2.02	02/24/2025	08:45
6	Q1413-01	1	Solid	20.02	20	23.2	8.37	02/24/2025	08:55
7	Q1413-01DUP	1	Solid	20.03	20	23.3	8.38	02/24/2025	08:56
8	Q1414-01	1	Solid	20.02	20	23.1	6.13	02/24/2025	09:00
9	CCV2	1	Water	NA	NA	20.3	12.02	02/24/2025	09:05

WORKLIST(Hardcopy Internal Chain)

LB 134783

WorkList Name : ph s q1414

WorkList ID : 187835

Department : Wet-Chemistry

Date : 02-24-2025 07:48:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1413-01	BU-2-022125	Solid	pH	Cool 4 deg C	PSEG05	H31	02/21/2025	9045D
Q1414-01	BUR-1316	Solid	pH	Cool 4 deg C	PSEG03	H31	02/21/2025	9045D

Date/Time 02/24/25 08:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 02/24/25 11:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]