

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134799

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/25/2025	15:45
2	CAL2	1	Water	NA	NA	20.2	7.00	02/25/2025	15:46
3	CAL3	1	Water	NA	NA	20.2	10.02	02/25/2025	15:50
4	ICV	1	Water	NA	NA	20.2	7.01	02/25/2025	15:55
5	CCV1	1	Water	NA	NA	20.2	2.01	02/25/2025	16:00
6	Q1425-01	1	Water	NA	NA	20.1	6.37	02/25/2025	16:05
7	Q1425-01DUP	1	Water	NA	NA	20.2	6.38	02/25/2025	16:06
8	Q1425-02	1	Water	NA	NA	20.7	6.33	02/25/2025	16:16
9	Q1425-04	1	Water	NA	NA	20.8	6.81	02/25/2025	16:20
10	Q1425-05	1	Water	NA	NA	20.5	6.78	02/25/2025	16:22
11	Q1425-07	1	Water	NA	NA	20.6	6.83	02/25/2025	16:25
12	Q1425-08	1	Water	NA	NA	20.5	6.66	02/25/2025	16:30
13	Q1425-09	1	Water	NA	NA	20.8	6.33	02/25/2025	16:33
14	Q1425-10	1	Water	NA	NA	20.1	6.39	02/25/2025	16:37
15	Q1425-12	1	Water	NA	NA	20.3	6.46	02/25/2025	16:40
16	CCV2	1	Water	NA	NA	20.1	12.02	02/25/2025	16:45
17	Q1425-13	1	Water	NA	NA	20.3	6.45	02/25/2025	16:50
18	Q1425-14	1	Water	NA	NA	20.3	6.34	02/25/2025	16:52
19	Q1426-01	1	Water	NA	NA	20.7	6.74	02/25/2025	17:00
20	CCV3	1	Water	NA	NA	20.3	2.01	02/25/2025	17:05

WORKLIST(Hardcopy Internal Chain)

134799

WorkList Name : ph q1425

WorkList ID : 187877

Department : Wet-Chemistry

Date : 02-25-2025 15:26:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1425-01	0217-A	Water	pH	Cool 4 deg C	PSEG03	H31	02/25/2025	9040C
Q1425-02	0217-B	Water	pH	Cool 4 deg C	PSEG03	H31	02/25/2025	9040C
Q1425-04	0212-A	Water	pH	Cool 4 deg C	PSEG03	H31	02/25/2025	9040C
Q1425-05	0212-B	Water	pH	Cool 4 deg C	PSEG03	H31	02/25/2025	9040C
Q1425-07	0212-C	Water	pH	Cool 4 deg C	PSEG03	H31	02/25/2025	9040C
Q1425-08	0214	Water	pH	Cool 4 deg C	PSEG03	H31	02/25/2025	9040C
Q1425-09	0219-A	Water	pH	Cool 4 deg C	PSEG03	H31	02/25/2025	9040C
Q1425-10	0219-B	Water	pH	Cool 4 deg C	PSEG03	H31	02/25/2025	9040C
Q1425-12	0224-A	Water	pH	Cool 4 deg C	PSEG03	H31	02/25/2025	9040C
Q1425-13	0224-B	Water	pH	Cool 4 deg C	PSEG03	H31	02/25/2025	9040C
Q1425-14	0224-C	Water	pH	Cool 4 deg C	PSEG03	H31	02/25/2025	9040C
Q1426-01	50453	Water	pH	Cool 4 deg C	PSEG03	H33	02/25/2025	9040C

Date/Time 02/25/25 15:25

Raw Sample Received by: JDCSM

Raw Sample Relinquished by: JDCSM

Date/Time 02/25/25

Raw Sample Received by: JDCSM

Raw Sample Relinquished by: JDCSM