

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB134800

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/25/2025	16:35
2	CAL2	1	Water	NA	NA	20.2	7.00	02/25/2025	16:36
3	CAL3	1	Water	NA	NA	20.2	10.02	02/25/2025	16:37
4	ICV	1	Water	NA	NA	20.3	7.01	02/25/2025	16:40
5	CCV1	1	Water	NA	NA	20.3	2.01	02/25/2025	16:45
6	Q1420-04	1	Solid	20.02	20	23.7	8.70	02/25/2025	16:50
7	Q1420-08	1	Solid	20.03	20	23.6	8.53	02/25/2025	17:10
8	Q1427-02	1	Solid	20.02	20	22.7	7.83	02/25/2025	17:15
9	Q1427-02DUP	1	Solid	20.03	20	22.8	7.84	02/25/2025	17:16
10	CCV2	1	Water	NA	NA	20.3	12.02	02/25/2025	17:20

WORKLIST(Hardcopy Internal Chain)

VB 134800

WorkList Name : corrosivity p1426

WorkList ID : 187878

Department : Wet-Chemistry

Date : 02-25-2025 15:27:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1420-04	TP-1-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	N41	02/24/2025	9045D
Q1420-08	TP-2-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	N41	02/24/2025	9045D
Q1427-02	VNJ-227	Solid	Corrosivity	Cool 4 deg C	PSEG03	H21	02/25/2025	9045D

Date/Time 02/25/25 15:45
 Raw Sample Received by: JB (WCI)
 Raw Sample Relinquished by: JB (WCI)

Date/Time 02/25/25 18:30
 Raw Sample Received by: JB (WCI)
 Raw Sample Relinquished by: JB (WCI)