

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB134814

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3159

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	02/26/2025	12:10
2	Q1408-01		1	15	100.00	>212.0	>212.0	02/26/2025	12:40
3	Q1408-01DUP		1	15	100.00	>212.0	>212.0	02/26/2025	13:10
4	Q1426-01		1	14	100.00	>212.0	>212.0	02/26/2025	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6134814

WorkList Name : fp-2-25

WorkList ID : 187859

Department : Wet-Chemistry

Date : 02-25-2025 08:16:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1408-01	813	Water	Flash Point	Cool 4 deg C	PSEG03	H23	02/21/2025	1010B
Q1426-01	50453	Water	Flash Point	Cool 4 deg C	PSEG03	H33	02/25/2025	1010B

Date/Time 02/26/2025 11:50
Raw Sample Received by: RM (wcy)
Raw Sample Relinquished by: MR (wcy)

Date/Time 02/26/2025 14:10
Raw Sample Received by: JPC (wcy)
Raw Sample Relinquished by: RM (wcy)