

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134827

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/27/2025	08:25
2	CAL2	1	Water	NA	NA	20.2	7.00	02/27/2025	08:26
3	CAL3	1	Water	NA	NA	20.2	10.02	02/27/2025	08:30
4	ICV	1	Water	NA	NA	20.2	7.00	02/27/2025	08:33
5	CCV1	1	Water	NA	NA	20.3	2.01	02/27/2025	08:37
6	Q1447-01	1	Solid	20.02	20	23.4	6.86	02/27/2025	08:45
7	Q1447-01DUP	1	Solid	20.03	20	23.5	6.87	02/27/2025	08:46
8	CCV2	1	Water	NA	NA	20.2	12.02	02/27/2025	08:50

WORKLIST(Hardcopy Internal Chain)

VB 134827

WorkList Name : ph s q1447

WorkList ID : 187913

Department : Wet-Chemistry

Date : 02-27-2025 08:12:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1447-01	WC1	Solid	pH	Cool 4 deg C	GENV01	H33	02/23/2025	9045D

Date/Time 02/27/25 08:15

Raw Sample Received by: SO WCC

Raw Sample Relinquished by: CP SM

Date/Time 02/27/25 12:30

Raw Sample Received by: CP SM

Raw Sample Relinquished by: SO WCC